

Invacare® Action 3 NG

en Manual Wheelchair User Manual



Contents

1	General	3
	1.1 Introduction	
	1.2 Symbols in this Document	
	1.3 Warranty Information	
	1.4 Limitation of Liability	
	1.5 Compliance	3
	1.5.1 Product-specific standards	. 3
	1.6 Service Life	
2	Safety	. 4
	2.1 Safety Information	. 4
	2.2 Safety devices	5
	2.3 Labels and symbols on the product	
3	Product Overview	. 5
	3.1 Product Description	. 5
	3.2 Intended Use	6
	3.3 Main parts of the wheelchair	
	3.4 Parking Brakes	
	-	
	3.5 Backrest	
	3.5.1 Standard Backrest Cover	
	3.5.2 Adjustable Backrest Cover	7
	3.5.3 Backrest cover height	. 7
	3.5.4 Fixed backrest	
	3.5.5 Folding backrest	
	3.5.6 Reclining Backrest	
	3.5.7 Angle adjustable backrest (-15°/ +15°)	
	3.5.8 Backbrace	
	3.5.9 Push handles	. 9
	3.6 Mudguard	. 10
	3.7 Armrests	10
	3 7 1 Tuhular armrest height-adjustable swiveling	10
	3.7.1 Tubular armrest, height-adjustable, swiveling	
	3.7.2 Armrest, stepless height adjustable, removable	.10
	3.7.2 Armrest, stepless height adjustable, removable3.7.3 Armrest, height adjustable, removable	.10 .11
	3.7.2 Armrest, stepless height adjustable, removable3.7.3 Armrest, height adjustable, removable3.7.4 Armrest, foldable and removable	.10 .11 .11
	 3.7.2 Armrest, stepless height adjustable, removable 3.7.3 Armrest, height adjustable, removable 3.7.4 Armrest, foldable and removable 3.7.5 Armrest, height adjustable, foldable and removable 	.10 .11 .11 .11
	3.7.2 Armrest, stepless height adjustable, removable 3.7.3 Armrest, height adjustable, removable 3.7.4 Armrest, foldable and removable 3.7.5 Armrest, height adjustable, foldable and removable 3.8 Leg rests	.10 .11 .11 .11
	 3.7.2 Armrest, stepless height adjustable, removable 3.7.3 Armrest, height adjustable, removable 3.7.4 Armrest, foldable and removable 3.7.5 Armrest, height adjustable, foldable and removable 	.10 .11 .11 .11
	3.7.2 Armrest, stepless height adjustable, removable 3.7.3 Armrest, height adjustable, removable 3.7.4 Armrest, foldable and removable 3.7.5 Armrest, height adjustable, foldable and removable 3.8 Leg rests	.10 .11 .11 .12 .12
	3.7.2 Armrest, stepless height adjustable, removable 3.7.3 Armrest, height adjustable, removable 3.7.4 Armrest, foldable and removable 3.7.5 Armrest, height adjustable, foldable and removable 3.8 Leg rests 3.8.1 Swing away leg rests	.10 .11 .11 .12 .12
	3.7.2 Armrest, stepless height adjustable, removable 3.7.3 Armrest, height adjustable, removable 3.7.4 Armrest, foldable and removable 3.7.5 Armrest, height adjustable, foldable and removable 3.8 Leg rests 3.8.1 Swing away leg rests 3.8.2 Swing away, Angle Adjustable Leg Rests 3.9 Anti-tipper	.10 .11 .11 .12 .12 .12
	3.7.2 Armrest, stepless height adjustable, removable 3.7.3 Armrest, height adjustable, removable 3.7.4 Armrest, foldable and removable 3.7.5 Armrest, height adjustable, foldable and removable 3.8 Leg rests 3.8.1 Swing away leg rests 3.8.2 Swing away, Angle Adjustable Leg Rests 3.9 Anti-tipper 3.10 Seat	.10 .11 .11 .12 .12 .12 .13 .14
	3.7.2 Armrest, stepless height adjustable, removable 3.7.3 Armrest, height adjustable, removable 3.7.4 Armrest, foldable and removable 3.7.5 Armrest, height adjustable, foldable and removable 3.8 Leg rests 3.8.1 Swing away leg rests 3.8.2 Swing away, Angle Adjustable Leg Rests 3.9 Anti-tipper 3.10 Seat 3.10.1 Adjusting the seat depth	.10 .11 .11 .12 .12 .12 .13 .14
	3.7.2 Armrest, stepless height adjustable, removable 3.7.3 Armrest, height adjustable, removable 3.7.4 Armrest, foldable and removable 3.7.5 Armrest, height adjustable, foldable and removable 3.8 Leg rests 3.8.1 Swing away leg rests 3.8.2 Swing away, Angle Adjustable Leg Rests 3.9 Anti-tipper 3.10 Seat 3.10.1 Adjusting the seat depth 3.10.2 Adjusting the shape of the seat	.10 .11 .11 .12 .12 .13 .14 .14
	3.7.2 Armrest, stepless height adjustable, removable 3.7.3 Armrest, height adjustable, removable 3.7.4 Armrest, foldable and removable 3.7.5 Armrest, height adjustable, foldable and removable 3.8 Leg rests 3.8.1 Swing away leg rests 3.8.2 Swing away, Angle Adjustable Leg Rests 3.9 Anti-tipper 3.10 Seat 3.10.1 Adjusting the seat depth	.10 .11 .11 .12 .12 .13 .14 .14
4	3.7.2 Armrest, stepless height adjustable, removable 3.7.3 Armrest, height adjustable, removable 3.7.4 Armrest, foldable and removable 3.7.5 Armrest, height adjustable, foldable and removable 3.8 Leg rests 3.8.1 Swing away leg rests 3.8.2 Swing away, Angle Adjustable Leg Rests 3.9 Anti-tipper 3.10 Seat 3.10.1 Adjusting the seat depth 3.10.2 Adjusting the shape of the seat 3.10.3 Seat cushion	.10 .11 .11 .12 .12 .13 .14 .14 .15
4	3.7.2 Armrest, stepless height adjustable, removable 3.7.3 Armrest, height adjustable, removable 3.7.4 Armrest, foldable and removable 3.7.5 Armrest, height adjustable, foldable and removable 3.8 Leg rests 3.8.1 Swing away leg rests 3.8.2 Swing away, Angle Adjustable Leg Rests 3.9 Anti-tipper 3.10 Seat 3.10.1 Adjusting the seat depth 3.10.2 Adjusting the shape of the seat 3.10.3 Seat cushion	.10 .11 .11 .12 .12 .13 .14 .14 .15
4	3.7.2 Armrest, stepless height adjustable, removable 3.7.3 Armrest, height adjustable, removable 3.7.4 Armrest, foldable and removable 3.7.5 Armrest, height adjustable, foldable and removable 3.8 Leg rests 3.8.1 Swing away leg rests 3.8.2 Swing away, Angle Adjustable Leg Rests 3.9 Anti-tipper 3.10 Seat 3.10.1 Adjusting the seat depth 3.10.2 Adjusting the shape of the seat 3.10.3 Seat cushion Options 4.1 Comfort Version	.10 .11 .11 .12 .12 .13 .14 .14 .15 .15
4	3.7.2 Armrest, stepless height adjustable, removable 3.7.3 Armrest, height adjustable, removable 3.7.4 Armrest, foldable and removable 3.7.5 Armrest, height adjustable, foldable and removable 3.8 Leg rests 3.8.1 Swing away leg rests 3.8.2 Swing away, Angle Adjustable Leg Rests 3.9 Anti-tipper 3.10 Seat 3.10.1 Adjusting the seat depth 3.10.2 Adjusting the shape of the seat 3.10.3 Seat cushion Options 4.1 Comfort Version 4.2 Transit version	.10 .11 .11 .12 .12 .13 .14 .14 .15 .15
4	3.7.2 Armrest, stepless height adjustable, removable 3.7.3 Armrest, height adjustable, removable 3.7.4 Armrest, foldable and removable 3.7.5 Armrest, height adjustable, foldable and removable 3.8 Leg rests 3.8.1 Swing away leg rests 3.8.2 Swing away, Angle Adjustable Leg Rests 3.9 Anti-tipper 3.10 Seat 3.10.1 Adjusting the seat depth 3.10.2 Adjusting the shape of the seat 3.10.3 Seat cushion Options 4.1 Comfort Version 4.2 Transit version 4.3 Dual Handrim DHR	.10 .11 .11 .12 .12 .13 .14 .14 .15 .15 .15
4	3.7.2 Armrest, stepless height adjustable, removable 3.7.3 Armrest, height adjustable, removable 3.7.4 Armrest, foldable and removable 3.7.5 Armrest, height adjustable, foldable and removable 3.8 Leg rests 3.8.1 Swing away leg rests 3.8.2 Swing away, Angle Adjustable Leg Rests 3.9 Anti-tipper 3.10 Seat 3.10.1 Adjusting the seat depth 3.10.2 Adjusting the shape of the seat 3.10.3 Seat cushion Options 4.1 Comfort Version 4.2 Transit version 4.3 Dual Handrim DHR 4.4 One arm drive by lever	.10 .11 .11 .12 .12 .13 .14 .14 .15 .15 .15
4	3.7.2 Armrest, stepless height adjustable, removable 3.7.3 Armrest, height adjustable, removable 3.7.4 Armrest, foldable and removable 3.7.5 Armrest, height adjustable, foldable and removable 3.8 Leg rests 3.8.1 Swing away leg rests 3.8.2 Swing away, Angle Adjustable Leg Rests 3.9 Anti-tipper 3.10 Seat 3.10.1 Adjusting the seat depth 3.10.2 Adjusting the shape of the seat 3.10.3 Seat cushion Options 4.1 Comfort Version 4.2 Transit version 4.3 Dual Handrim DHR	.10 .11 .11 .12 .12 .13 .14 .14 .15 .15 .15 .15
4	3.7.2 Armrest, stepless height adjustable, removable 3.7.3 Armrest, height adjustable, removable 3.7.4 Armrest, foldable and removable 3.7.5 Armrest, height adjustable, foldable and removable 3.8 Leg rests 3.8.1 Swing away leg rests 3.8.2 Swing away, Angle Adjustable Leg Rests 3.9 Anti-tipper 3.10 Seat 3.10.1 Adjusting the seat depth 3.10.2 Adjusting the shape of the seat 3.10.3 Seat cushion Options 4.1 Comfort Version 4.2 Transit version 4.3 Dual Handrim DHR 4.4 One arm drive by lever	.10 .11 .11 .12 .12 .13 .14 .14 .15 .15 .15 .15 .17
4	3.7.2 Armrest, stepless height adjustable, removable 3.7.3 Armrest, height adjustable, removable 3.7.4 Armrest, foldable and removable 3.7.5 Armrest, height adjustable, foldable and removable 3.8 Leg rests 3.8.1 Swing away leg rests 3.8.2 Swing away, Angle Adjustable Leg Rests 3.9 Anti-tipper 3.10 Seat 3.10.1 Adjusting the seat depth 3.10.2 Adjusting the shape of the seat 3.10.3 Seat cushion Options 4.1 Comfort Version 4.2 Transit version 4.3 Dual Handrim DHR 4.4 One arm drive by lever 4.5 Posture belt	.10 .11 .12 .12 .13 .14 .14 .15 .15 .15 .16 .17
4	3.7.2 Armrest, stepless height adjustable, removable 3.7.3 Armrest, height adjustable, removable 3.7.4 Armrest, foldable and removable 3.7.5 Armrest, height adjustable, foldable and removable 3.8.1 Swing away leg rests 3.8.2 Swing away, Angle Adjustable Leg Rests 3.9 Anti-tipper 3.10 Seat 3.10.1 Adjusting the seat depth 3.10.2 Adjusting the shape of the seat 3.10.3 Seat cushion Options 4.1 Comfort Version 4.2 Transit version 4.3 Dual Handrim DHR 4.4 One arm drive by lever 4.5 Posture belt 4.6 Headrest 4.7 Lateral positioning pads (for adjustable backrest cover only)	.10 .11 .12 .12 .13 .14 .15 .15 .15 .16 .17 .18
4	3.7.2 Armrest, stepless height adjustable, removable 3.7.3 Armrest, height adjustable, removable 3.7.4 Armrest, foldable and removable 3.7.5 Armrest, height adjustable, foldable and removable 3.8.1 Swing away leg rests 3.8.1 Swing away, Angle Adjustable Leg Rests 3.8.2 Swing away, Angle Adjustable Leg Rests 3.9 Anti-tipper 3.10 Seat 3.10.1 Adjusting the seat depth 3.10.2 Adjusting the shape of the seat 3.10.3 Seat cushion Options 4.1 Comfort Version 4.2 Transit version 4.3 Dual Handrim DHR 4.4 One arm drive by lever 4.5 Posture belt 4.6 Headrest 4.7 Lateral positioning pads (for adjustable backrest cover only) 4.8 Tipper aid	.10 .11 .12 .12 .13 .14 .14 .15 .15 .15 .16 .17 .18 .18 .19
4	3.7.2 Armrest, stepless height adjustable, removable 3.7.3 Armrest, height adjustable, removable 3.7.4 Armrest, foldable and removable 3.7.5 Armrest, height adjustable, foldable and removable 3.8.1 Swing away leg rests 3.8.1 Swing away, Angle Adjustable Leg Rests 3.8.2 Swing away, Angle Adjustable Leg Rests 3.9 Anti-tipper 3.10 Seat 3.10.1 Adjusting the seat depth 3.10.2 Adjusting the shape of the seat 3.10.3 Seat cushion 4.1 Comfort Version 4.2 Transit version 4.3 Dual Handrim DHR 4.4 One arm drive by lever 4.5 Posture belt 4.6 Headrest 4.7 Lateral positioning pads (for adjustable backrest cover only) 4.8 Tipper aid 4.9 Table tray	.10 .11 .12 .12 .13 .14 .14 .15 .15 .15 .16 .17 .17 .18 .19 .19
4	3.7.2 Armrest, stepless height adjustable, removable 3.7.3 Armrest, height adjustable, removable 3.7.4 Armrest, foldable and removable 3.7.5 Armrest, height adjustable, foldable and removable 3.8.1 Swing away leg rests 3.8.1 Swing away, Angle Adjustable Leg Rests 3.8.2 Swing away, Angle Adjustable Leg Rests 3.9 Anti-tipper 3.10 Seat 3.10.1 Adjusting the seat depth 3.10.2 Adjusting the shape of the seat 3.10.3 Seat cushion 4.0 Options 4.1 Comfort Version 4.2 Transit version 4.3 Dual Handrim DHR 4.4 One arm drive by lever 4.5 Posture belt 4.6 Headrest 4.7 Lateral positioning pads (for adjustable backrest cover only) 4.8 Tipper aid 4.9 Table tray 4.10 Half tray	.10 .11 .11 .12 .12 .13 .14 .15 .15 .15 .16 .17 .17 .18 .19 .19
4	3.7.2 Armrest, stepless height adjustable, removable 3.7.3 Armrest, height adjustable, removable 3.7.4 Armrest, foldable and removable 3.7.5 Armrest, height adjustable, foldable and removable 3.8.1 Swing away leg rests 3.8.1 Swing away, Angle Adjustable Leg Rests 3.8.2 Swing away, Angle Adjustable Leg Rests 3.9 Anti-tipper 3.10 Seat 3.10.1 Adjusting the seat depth 3.10.2 Adjusting the shape of the seat 3.10.3 Seat cushion 4.0 Options 4.1 Comfort Version 4.2 Transit version 4.3 Dual Handrim DHR 4.4 One arm drive by lever 4.5 Posture belt 4.6 Headrest 4.7 Lateral positioning pads (for adjustable backrest cover only) 4.8 Tipper aid 4.9 Table tray 4.10 Half tray 4.11 Add a half tray cushion	.10 .11 .11 .12 .12 .13 .14 .15 .15 .15 .16 .17 .17 .18 .19 .19 .19 .20
4	3.7.2 Armrest, stepless height adjustable, removable 3.7.3 Armrest, height adjustable, removable 3.7.4 Armrest, foldable and removable 3.7.5 Armrest, height adjustable, foldable and removable 3.8.1 Swing away leg rests 3.8.1 Swing away, Angle Adjustable Leg Rests 3.8.2 Swing away, Angle Adjustable Leg Rests 3.9 Anti-tipper 3.10 Seat 3.10.1 Adjusting the seat depth 3.10.2 Adjusting the shape of the seat 3.10.3 Seat cushion 4.0 Options 4.1 Comfort Version 4.2 Transit version 4.3 Dual Handrim DHR 4.4 One arm drive by lever 4.5 Posture belt 4.6 Headrest 4.7 Lateral positioning pads (for adjustable backrest cover only) 4.8 Tipper aid 4.9 Table tray 4.10 Half tray 4.11 Add a half tray cushion 4.12 Drip stand	.10 .11 .11 .12 .12 .13 .14 .14 .15 .15 .15 .15 .16 .17 .17 .18 .19 .19 .20 .20
4	3.7.2 Armrest, stepless height adjustable, removable 3.7.3 Armrest, height adjustable, removable 3.7.4 Armrest, foldable and removable 3.7.5 Armrest, height adjustable, foldable and removable 3.8.1 Swing away leg rests 3.8.1 Swing away, Angle Adjustable Leg Rests 3.8.2 Swing away, Angle Adjustable Leg Rests 3.9 Anti-tipper 3.10 Seat 3.10.1 Adjusting the seat depth 3.10.2 Adjusting the shape of the seat 3.10.3 Seat cushion 4.0 Options 4.1 Comfort Version 4.2 Transit version 4.3 Dual Handrim DHR 4.4 One arm drive by lever 4.5 Posture belt 4.6 Headrest 4.7 Lateral positioning pads (for adjustable backrest cover only) 4.8 Tipper aid 4.9 Table tray 4.10 Half tray 4.11 Add a half tray cushion	.10 .11 .11 .12 .12 .13 .14 .14 .15 .15 .15 .15 .16 .17 .17 .18 .19 .19 .20 .20
4	3.7.2 Armrest, stepless height adjustable, removable 3.7.3 Armrest, height adjustable, removable 3.7.4 Armrest, foldable and removable 3.7.5 Armrest, height adjustable, foldable and removable 3.8.1 Swing away leg rests 3.8.1 Swing away, Angle Adjustable Leg Rests 3.8.2 Swing away, Angle Adjustable Leg Rests 3.9 Anti-tipper 3.10 Seat 3.10.1 Adjusting the seat depth 3.10.2 Adjusting the shape of the seat 3.10.3 Seat cushion 4.0 Options 4.1 Comfort Version 4.2 Transit version 4.3 Dual Handrim DHR 4.4 One arm drive by lever 4.5 Posture belt 4.6 Headrest 4.7 Lateral positioning pads (for adjustable backrest cover only) 4.8 Tipper aid 4.9 Table tray 4.10 Half tray 4.11 Add a half tray cushion 4.12 Drip stand	.10 .11 .11 .12 .12 .13 .14 .14 .15 .15 .15 .16 .17 .17 .18 .19 .19 .20 .20

	4.15 Passive illumination	. 21
	4.16 Cane holder	. 21
5	S Setup	21
	5.1 Safety Information	
	5.2 Delivery Check	.21
6	Using the wheelchair	21
	6.1 Safety Information	
	6.2 Braking during use	
	6.3 Move to/from the wheelchair	. 22
	6.4 Driving and Steering the Wheelchair	
	6.5 Negociating steps and kerbs	
	6.6 Going up and down stairs	
	6.7 Negotiating Ramps and Slopes	
	6.8 Stability and balance when seated	24
7	7 Transport	
	7.1 Safety Information	
	7.2 Folding and unfolding the wheelchair	
	7.3 Lifting the wheelchair	
	7.4 Removing and fitting the rear wheels	
	7.5 Transporting the Manual Wheelchair without Occupant7.6 Transporting the Occupied Wheelchair in a Vehicle	
8		
	8.1 Safety Information	
	8.2 Maintenance schedule	
	8.3. Cleaning and disinfection	
	8.3.2 Cleaning Intervals	
	8.3.3 Cleaning	
	8.3.4 Washing	
	8.3.5 Disinfection	
_	Troubleshooting	
7	9.1 Safety Information	
	9.2 Identifying and repairing faults	
_		
1	LO After use	
	10.1 Storage	
	10.2 Disposal	
1	11 Technical data	
	11.1 Dimensions and weight	
	11.2 Maximum Weight of Removable Parts	
	11.3 Tyres	
	11.5 Environmental conditions	

© 2025 Invacare International GmbH

All rights reserved. Republication, duplication or modification in whole or in part is prohibited without prior written permission from Invacare. Trademarks are identified by ™ and ®. All trademarks are owned by or licensed to Invacare International GmbH or its subsidiaries unless otherwise noted. Invacare reserves the right to alter product specifications without further notice.

1 General

1.1 Introduction

This user manual contains important information about the handling of the product. To ensure safety when using the product, read the user manual carefully and follow the safety instructions.

Only use this product if you have read and understood this manual. Seek additional advice from a healthcare professional who is familiar with your medical condition and clarify any questions regarding the correct use and necessary adjustment with the healthcare professional.

Note that there may be sections in this document, which are not relevant to your product, since this document applies to all available models (on the date of printing). If not otherwise stated, each section in this document refers to all models of the product.

The models and configurations available in your country can be found in the country-specific sales documents. Invacare reserves the right to alter product specifications without further notice.

Previous product versions may not be described in this manual's current revision. If you require assistance, please contact Invacare.

Before reading this document, make sure you have the latest version. You find the latest version as a PDF on the Invacare website.

If you find that the font size in the printed document is difficult to read, you can download the PDF version from the website. The PDF can then be scaled on screen to a font size that is more comfortable for you.

For more information about the product, for example product safety notices and product recalls, contact your Invacare distributor. See addresses at the end of this document.

In case of a serious incident with the product, you should inform the manufacturer and the competent authority in your country.

1.2 Symbols in this Document

Symbols and signal words are used in this document and apply to hazards or unsafe practices which could result in personal injury or property damage. This document is printed in greyscale. For your information, the safety messages have the following colour coding according to ANSI Z535.6: Danger (Red), Warning (Orange), Caution (Yellow) and Notice (Blue). See the information below for definitions of the signal words.



WARNING!

Indicates a hazardous situation that could result in serious injury or death if it is not avoided.



CAUTION!

Indicates a hazardous situation that could result in minor or slight injury if it is not avoided.



NOTICE!

Indicates a hazardous situation that could result in damage to property if it is not avoided.



Tips and Recommendations

Gives useful tips, recommendations, and information for efficient, trouble-free use.



Triman

Indicates recycling and sorting rules (only relevant for France).



UKRP

Indicates if a product is not manufactured in the UK.

1.3 Warranty Information

We provide a manufacturer's warranty for the product in accordance with our General Terms and Condition of Business in the respective countries.

Warranty claims can only be made through the provider from whom the product was obtained

1.4 Limitation of Liability

Invacare accepts no liability for damage arising from:

- · Non-compliance with the user manual
- · Incorrect use
- Natural wear and tear
- Incorrect assembly or set-up by the purchaser or a third party
- Technical modifications
- Unauthorised modifications and/or use of unsuitable spare parts

1.5 Compliance

Quality is fundamental to the company's operation, working within the disciplines of ISO 13485. This product features the CE mark, in compliance with the Medical Device Regulation 2017/745 Class I.

This product features the UKCA mark, in compliance with Part II UK MDR 2002 (as amended) Class I.

We are continuously working towards ensuring that the company's impact on the environment, locally and globally, is reduced to a minimum.

We only use REACH compliant materials and components

1.5.1 Product-specific standards

The wheelchair has been tested in accordance with EN 12183. It includes testing for flammability. For further information about local standards and regulations, contact your local Invacare representative. See addresses at the end of this document.

1.6 Service Life

The expected service life of this product is five years when used daily and in accordance with the safety instructions, maintenance intervals and correct use, stated in this manual. The effective service life can vary according to frequency and intensity of use.

2 Safety

2.1 Safety Information

This section contains important safety information for the protection of the wheelchair occupant and assistant and for safe, trouble-free use of the wheelchair.



WARNING!

Risk of death or serious injury

In case of fire or smoke, wheelchair occupants are at particular risk of death or serious injury, when they are not able to move away from the source of fire or smoke. Lighted matches, lighter and cigarettes can cause an open flame in the wheelchair surroundings or on clothes.

- Avoid using or storing the wheelchair near open flames or combustible products.
- Do not smoke while using the wheelchair.



WARNING!

Risk of accidents and serious injury

Accidents with resulting serious injury can occur if the wheelchair is wrongly adjusted.

 Adjustments at the wheelchair always must be carried out by a qualified technician.



WARNING!

Risk of overturning

The longitudinal position of the rear wheels axis of the wheelchair compared to the backrest position can affect its stability..

- A forwards position makes the wheelchair less stable and increase the risk of tipping backwards, but improves its maneuverability by a better grip position of the handrim and a short turning radius.
- Conversely, by moving the rear wheels axis backwards, the wheelchair is more stable and tilts less easily, but its maneuverability is reduced.
- Depending on the user's abilities and its particular safety limits, the decrease in stability can be compensated for by installing an antitipper device.



WARNING! Risk of tilting

The position of the rear wheel axle and the angle of your wheelchair's backrest are two of the key adjustments that can affect your stability whilst sitting in the wheelchair.

 These changes, any front wheels position and/or angle fork adjustments to the set-up of your wheelchair must only be performed at the direction of a qualified technician who has fully risk assessed the implication of any changes to the wheelchair configuration. Therefore, consult with your authorized provider.



WARNING!

Risk due to driving style being unsuitable for the conditions

There is a risk of skidding on wet ground, gravel or uneven terrain.

 Always adjust your speed and driving style to the conditions (weather, surface, individual ability, etc.).



WARNING!

Risk of injury

In a collision you could sustain injury to parts of your body that extend beyond the wheelchair (e.g. feet or hands).

- Avoid an unbraked collision.
- Never drive into an object head-on.
- Drive carefully through narrow passages.



WARNING!

Riskdue to wheelchair being out of control

At high speed you could lose control of your wheelchair and overturn.

- Never exceed a speed of 7 km/h.
- Avoid collisions in general.



CAUTION!

Risk of burning

The wheelchair components can heat up when exposed to external sources of heat.

- Do not expose the wheelchair to strong sunlight before use.
- Before usage, check all components that come into contact with your skin for their temperature.



CAUTION!

Risk of getting fingers caught

There is always a risk of getting, e.g. fingers or arms, caught in the moving parts of the wheelchair.

 Make sure when activating the folding or insertion mechanisms of moving parts, such as the removable axle of the rear wheel, folding backrest or anti-tipper, that nothing becomes caught.



CAUTION!

Risk of tipping

The hanging of additional load (back pack or similar items) onto your chair backposts can affect the rearward stability of your wheelchair, especially when used in combination with angle adjustable backrests.

 Invacare strongly recommend the use of antitipper (available as an option) when using your backposts with additional load.



CAUTION! Risk of injury

 In case of pressure sore or injured skin, protect your injure to avoid a direct contact with the fabrics of the device. Refer to a healthcare professional for medical advices.

1608467-I



CAUTION! Risk of injury

When combining the wheelchair with another device, the restrictions of both devices apply for the combination. E.g. the maximum user weight of the device may be lower.

- Only use combinations with other devices which are allowed by Invacare. Contact your authorized provider for more information.
- Before use, read the user manual of each device and check the restrictions.

2.2 Safety devices



WARNING! Risk of accidents

Safety devices that are incorrectly set or no longer working (brakes, anti-tipper) can cause accidents.

Always check that the safety devices are working before you use the wheelchair and have them checked regularly by a qualified technician or your authorized provider.



CAUTION!

Risk of Injury


Non-original or wrong parts may affect the function and safety of the product.

- Only use original parts for the product in use.
- Due to regional differences, refer to your local Invacare catalogue or website for available options or contact your Invacare distributor. See addresses at the end of this document.

The functions of the safety devices are described in 3 Product Overview, page 5.

2.3 Labels and symbols on the product

The identification label is attached to the frame of your wheelchair and gives the following information:



- (A) Manufacturer's name and address
- Maximum load of the wheelchair (B)
- European conformity mark (C)
- **D** United Kingdom conformity mark
- (E) Consult user manual
- (F)Medical device
- G **UDI** barcode
- \oplus Unique Device Identifier with bar code
- Serial Number (I)
- (]) Reference number
- $\widehat{(K)}$ Seat width
- (L) Product name
- Date of manufacture

Maximum load of the wheelchair or Maximum load or wheelchair or wheelc	
Consult user manual	
\triangle	Indicates a hazardous situation that could result

Non-locking side parts warning label



Do not lift the wheelchair by non-locking side parts.

in serious injury or death if it is not avoided.

Posture belt information label



Do not lift the wheelchair by non-locking side

Reclining mechanical backrest information label



Push on the back support tubes (1) before operating the levers (2), this is to release the auto-locked security system. This label is attached to the upper part of the backrest cane (right side).

Snap hook symbols

Depending on the configuration, some wheelchairs may be used as a seat in a motor vehicle, some may not.



Tie-down positions where the restraint system straps must be placed in case of transporting the occupied wheelchair in a motor vehicle. This symbol is only attached to the wheelchair when it is ordered with the transportation kit option.

Symbol for wheelchair not intended to be used as a seat in a motor vehicle.

This wheelchair is not configured for passenger transport in a motor vehicle. This symbol is attached to the frame close to the identification label.





WARNING!

Risk of serious injury or death Depending on the configuration, some wheelchairs may be used as a seat in a motor vehicle, some may not.

 Do not use the wheelchair for seat in a motor vehicle in.

Product Overview

3.1 Product Description

This is a manual wheelchair with crossbar folding mechanism and swing-away leg rests.

1608467-I

NOTICE!

The wheelchair is manufactured and configured individually to the specifications in the order. The specification must be performed by a healthcare professional/ clinician according to the occupant's requirements and health condition.

- Consult a healthcare professional if you intend to adapt the wheelchair configuration.
- Any adaptation should be performed by a qualified technician.

3.2 Intended Use

The medium active wheelchair is intended to provide mobility to persons limited to a sitting position, who are propelling the wheelchair themselves frequently or use the wheelchair for assisted transport.

The wheelchair is intended for persons ages 12 and up (adolescents and adults). The weight of the wheelchair occupant must not exceed the maximum occupant mass as stated in the Technical Data section and on the identification label.

The wheelchair may be used indoors and outdoors on level ground and accessible terrain.

Intended users

The wheelchair is intended for persons ages 12 and up (adolescents and adults). The weight of the wheelchair occupant must not exceed the maximum user weight as stated in the Technical Data section and on the identification label. The intended user is the wheelchair occupant and/or an assistant.

The user should physically and mentally be able to use the wheelchair safely (e.g. to propel, steer, brake).


Indications

- Considerable to complete impairment of mobility due to structural and/ or functional damage to the lower extremities.
- Sufficient strength and gripping function of arms and hands to propel the wheelchair.

Contraindications

There are no contra-indications known, when using the wheelchair as intended.

3.3 Main parts of the wheelchair



- A Push handle
- B Backrest

- © Rear wheel with handrim
- D Parking brake
- **E** Front fork with castor
- F Leg rests, swivelling
- **©** Footrest
- ⊕ Frame
- ① Seat
- ① Armrest
 - The equipment of your wheelchair may differ from the diagram as each wheelchair is manufactured individually to the specifications in the order.

3.4 Parking Brakes

The parking brakes are used to immobilize the wheelchair when it is stationary to prevent it from rolling away.



WARNING!

Risk of overturning if you brake sharply

If you engage the parking brakes while you are moving, the direction of movement can become uncontrollable and the wheelchair may stop suddenly, which can lead to a collision or to you falling out

 Never engage the parking brakes while you are moving.



WARNING!

Risk due to wheelchair being out of control

- The parking brakes must be operated simultaneously.
- Do not engage the parking brakes to slow down the wheelchair.
- Do not lean on the parking brakes for support or transfer.



WARNING!

Risk of overturning

 The parking brakes will not operate correctly unless there is sufficient air in the tyres. Ensure the correct tyre pressure, 11.3 Tyres, page 33.



CAUTION!


Risk of pinching or crushing

There may be a very small gap between the rear wheel or the armrest front socket and the parking brake with the risk that you could trap your fingers.

- Keep your fingers away from movable parts when using the brake, always keep your hand on the brake lever.
- The distance between the brake shoe and the tyre can be adjusted. The adjustment must be carried out by a qualified technician.


1608467-І

Standard brake



- 1. To engage the brake, push the brake lever forwards with the palm as far as possible.
- 2. To disengage the brake, pull the brake lever backwards with the low fingers.
 - The lever of the push handle can be folded back to facilitate transfers. To do so, pull up the lever and fold it backwards.
 - An assistant's brake (drum brake) is available as an option, the assistant can brake the wheelchair while it is moving. Please see section 6.2 Braking during use, page 21.

Pull brake



- 1. To engage the brake, pull the brake lever backwards with the low fingers as far as possible.
- 2. To disengage the brake, push the brake lever forwards with the palm.
 - The lever of the push handle can be folded back to facilitate transfers. To do so, pull up the lever and fold it backwards.

One arm drive push brake (available as an option)

This push brake allows the user to immobilise the wheelchair with a single arm (One brake lever; right or left hand).

3.5 Backrest

There are three types of backrest (Fixed, Folding and Reclining) available.



WARNING!

Risk of injury to the assistant and damage to the wheelchair

Tilting the chair with a heavy occupant can hurt the assistant's back and damage the chair.

 Make sure to be able to safely control the wheelchair with a heavy occupant before performing a tilting maneuver.

3.5.1 Standard Backrest Cover

A standard backrest cover is available for all backrest types. These covers are not adjustable.



WARNING! Risk of tipping


If the standard backrest cover become slack, the tipping point of your wheelchair is changed for the worse.

 Slack standard backrest cover must be replaced by a qualified technician.

3.5.2 Adjustable Backrest Cover

You can adjust the tension of the backrest cover as required.

Adjusting the backrest cover



- 1. Remove the backrest cushion (A).
- 2. Loosen the hook and loop strips ® on the rear of the backrest cover by simply pulling on them.
- 3. Tighten or loosen the strips as required then reattach them.



WARNING!

Risk of tipping

If the bands are set very loose, the tipping point of your wheelchair is changed for the worse.

Ensure that the bands are set correctly.



NOTICE!

 Do not overtighten the bands; this is to ensure that the geometry of the wheelchair is not changed.



NOTICE!

Only ever tighten the bands when the wheelchair is unfolded.

3.5.3 Backrest cover height

The backrest cover height can be adjusted for all backrest types.



WARNING! Risk of tipping

The hanging of additional load (back pack or similar items) onto your wheelchair back posts can affect the rearward stability of your wheelchair. This can cause the chair to tip backwards causing injury.

- Therefore, Invacare strongly recommend the use of anti-tip devices (available as an option) when using your back posts with additional load (back pack or similar items).
- The height of the backrest cover can be adjusted on the back posts. The adjustment must be carried out by a qualified technician.


1608467-I 7

3.5.4 Fixed backrest

Fixed backrest of 0° and 7°: these two backrests do not require adjustments.

3.5.5 Folding backrest

To save space for transporting the wheelchair, the backrest can be folded.




Folding and unfolding the backrest

- 2. To return to the initial position, bring the top in the vertical position; it locks automatically.

3.5.6 Reclining Backrest

To achieve a more comfort position for the user, the backrest can be reclined.





WARNING!

Risk of tipping

Do not move the wheelchair when the backrest is in reclined position.

 Invacare strongly recommend the use of anti-tip devices (available as an option) when the backrest is in reclined position.

Backrest angle

The angle of the backrest can be adjusted in four steps (0° to 30°) with the mechanical version, and step less with the gas struts version.



WARNING!

Risk of injury

- Always make these adjustments before the user settles down in the wheelchair to prevent injuries.
- Invacare strongly recommends the use of antitip devices (available as an option) in combination with all rear wheel positions when you use your backrest with an angle of 12° and more.



CAUTION!

Risk of pinching or crushing fingers

There is always a risk of getting fingers pinched in mechanical parts of the wheelchair.

 Be careful when activating the control levers of the reclining backrest.




NOTICE!

- Push on the backrest canes before operating the levers, this is to release the auto-locked security system.
- Release the auto-locked system by pushing on the backrest canes.
- 2. Simultaneously pull the control levers (A) to provides the same angle on both sides.
- 3. Release the levers when you reach the desired angle.

3.5.7 Angle adjustable backrest (-15°/ +15°)

To achieve a more comfort position for the user, the backrest can be reclined.



- Adjust the backrest angle thanks to the adjustable screw
- 2. Turn 5 mm Allen key clockwise to obtain a negative angle.
- 3. Alternatively, turn 5 mm Allen key under clockwise to obtain a positive angle.
- 4. Ensure both adjustments to the desired angle are equal on both sides.

Invacare recommends that the backrest angle adjustment is carried out by a qualified technician.



WARNING!

Risk of tipping

Do not move the wheelchair when the backrest is in reclined position.

 Invacare strongly recommend the use of anti-tip devices (available as an option) when the backrest is in reclined position.



WARNING! Risk of injury

Do not move the wheelchair when the backrest is in reclined position.

 Invacare strongly recommends the use of antitippers (available as an option) in combination with all rear wheel positions when you use your backrest with an angle of 12° and more.

1608467-І



CAUTION!

Risk of uncomfortable posture

An angle less than 90° between the seat and the backrest is uncomfortable for certain users.

 This adjustment must be performed by a qualified technician upon agreement by a prescribing physician. Please consult with your provider.

3.5.8 Backbrace

There are two types of backbraces (Standard backbrace and Folding backbrace) available.

Standard backbrace for the backrest


The backbrace links the two push handles, provides tension to the backrest upholstery and must always be in position on recliner versions.



CAUTION!

- Do not use the backbrace to push the wheelchair. The backbrace is not a push bar!
- Do not attempt to lift the wheelchair using the backbrace; it may come loose and it could break.
- Keep your fingers away from movable parts and sharp edges to prevent injuries.

Folding the backbrace



- Slightly loosen the two hand screws (a), pull up the backbrace (b) and swivel it downwards along the backrest (on right or left hand).
- 2. Fold the backbrace ® until it is in vertical position.

Unfolding the backbrace

- 1. Fold the backbrace ® upwards.
- 2. Align the backbrace ® with the push handle and push it down on the hand screw ⓐ.
- 3. Firmly tighten the hand screw (A) and makes sure that the TWO hand screws (A) are properly tightened .
 - lt is important to ensure that the backbrace is correctly fitted and engaged at all times when the chair is in use. It should not be left hanging loose (see label C on the backbrace).



Folding backbrace for the backrest

An optional folding backbrace is available to increase the rigidity of the backrest on wheelchairs with large back heights or widths.



CAUTION!

 Do not use the folding backbrace to lift or push the wheelchair. It could break.

Folding the backbrace



- Unlock the folding backbrace by pushing the joint A slightly downwards.
- 2. Fold the backbrace downwards.
 - The folding backbrace will automatically release when the chair is folded.

Unfolding the backbrace

- 1. Fold the backbrace upwards.
- Lock the folding backbrace by pulling the joint A slightly upwards.

3.5.9 Push handles




NOTICE!

 Always check the push handles before using the wheelchair, as to whether the hand grips are secure, cannot be turned and cannot be pulled off.

Height adjustable, rear set push handles (optional)

The height adjustable, rear set push handles allow any assistant to adjust the handles to a comfortable level.



 To set the height of the push handles, loosen the knob (A), push the handle to the desired position and tighten the knob again.




CAUTION

If the knob is not correctly tightened, the push handle could accidentally come off the backrest tube while pushing it upwards.

Make sure that the knob is firmly tightened.

Height adjustable, integrated push handles (optional)

The height adjustable, integrated push handles allow any assistant to adjust the handles to a comfortable level.



 To set the height of the push handles, loosen the knob (A), push the handle to the desired position and tighten the knob again.




CAUTION!

If the knob is not correctly tightened, the push handle could accidentally come off the backrest tube while pushing it upwards.

Make sure that the knob is firmly tightened.

3.6 Mudguard

In order to prevent dirt from the wheels spraying upwards, a removable mudguard can be fitted. The position of the mudguard can be adjusted.



Removing

- 1. Pull the mudguard upwards out of the holder.
- To adjust how easy or difficult it is to pull the mudguard out of the holder, change the degree to which the screws
 are tightened.

Fitting

1. Push the mudguard into the holder.

Adjusting the height

- 1. Loosen the screw (B) with 10 mm spanner in the groove (C) of the armrest tube.
- 2. Move the armrest upwards or downwards until you reach the desired height, while keeping screw (B) in place (at the top edge of the holder).
- 3. Tighten the screw ® (5–6 Nm).
- 4. Perform the setting on both sides.

3.7 Armrests



WARNING! Risk of injury

The side rests are not locked and can be easily pulled out upwards.

- Do not lift or transfer the wheelchair using the side rests.
- Do not use the side rests for transportation when carrying the wheelchair up- or downstairs.




CAUTION!

Risk of Pinching

 Keep your fingers away from movable parts during removing, fitting or adjusting the armrest.

3.7.1 Tubular armrest, height-adjustable, swiveling



Removing

 Raise the tubular armrest and pull it upwards out of the holder.

Fitting

1. Push the tubular armrest into the holder.


Adjusting the height

- 1. Pull the tubular armrest out of the holder.
- 2. Loosen the screw on the tube of the armrest and screw it back in at the desired height.
- 3. Push the tubular armrest back into the holder.
- 4. Perform the setting on both sides.

Swiveling

1. Raise the armrest a little and swivel it outwards.

3.7.2 Armrest, stepless height adjustable, removable



Removing

- 1. Apply the parking brake on both sides.
- 2. Previously press on the locking mechanism (a) located at the bottom of the armrest vertical bracket (if so equipped).
- 3. Pull the armrest by the arm cushion out of the holder.
- 4. To adjust how easy or difficult it is to pull the armrest out of the holder, change the degree to which the screws D are tightened.


Fitting

1. Push the armrest into the holder.

Adjusting the height

- 2. Move the armrest upwards or downwards until you reach the desired height, while keeping screw (B) in place (at the top edge of the holder).
- 3. Tighten the screw ® (5–6 Nm).
- 4. Perform the setting on both sides.

3.7.3 Armrest, height adjustable, removable



Removing

- 1. Previously press on the locking mechanism A located at the bottom of the armrest vertical bracket (if so equipped).
- 2. Pull the armrest by the arm cushion out of the holder.

Fitting

1. Push the armrest into the holder.

Adjusting the armpad height

- 2. Move the armrest upwards or downwards until you reach the desired height, while keeping screw (B) in place.
- 3. Tighten the screw ® (5–6 Nm).
- 4. Perform the setting on both sides.


Adjusting the armpad depth

- 1. Loosen the 2 screws $\ \ \ \$ below the armrest bracket.
- 2. Move the armrest frontwards or rearwards until you reach the desired depth, while keeping screw © in place.
- 3. Tighten the 2 screws © (1,5–2 Nm).
- 4. Perform the setting on both sides.

Adjusting the side protection height

- 1. Loosen the 2 screws of the side protection clamps ①.
- 2. Move the side protection upwards or downwards until you reach the desired height.
- 4. Perform the setting on both sides.

3.7.4 Armrest, foldable and removable



Folding

1. Grip the armrest by the armpad or by the front curve and fold it back.

Front removing

- Remove armrest from armrest socket by pulling straight up and fold it back.

Front fitting

- Push the armrest down and until the front release button
 snaps into its armrest socket hole.
- 2. Ensure button protrudes fully through socket hole.

Rear removing


1. Remove armrest from armrest bracket

by pulling straight up on the armrest.

Rear fitting

- 1. Position armrest support © over rear armrest bracket ©.
- 2. Push the armrest down until a clear "click" is heard when it's locked into the armrest bracket ©.

3.7.5 Armrest, height adjustable, foldable and removable



Folding

1. Grip the armrest by the armpad or by the front curve and fold it back.

Front removing

- Press and hold the armrest release push pin (A) at the front
- 2. Remove armrest from armrest socket by pulling straight up and fold it back.

Front fitting

- 1. Push the armrest down and until the front release button (B) snaps into its armrest socket hole.
- 2. Ensure button protrudes fully through socket hole.

Rear removing

1. Remove armrest from armrest bracket

by pulling straight up on the armrest.

Rear fitting

- 1. Position armrest support © over rear armrest bracket ©.
- 2. Push the armrest down until a clear "click" is heard when it's locked into the armrest bracket .

Adjusting the height

- 1. After loosening the fixing screw © with 5 mm Allen key or alternatively the thumbscrew, if present.
- 2. Adjust the height and tighten the fixing screw (£) (5 Nm) or firmly tighten the thumbscrew, into the desired position.
- 3. Perform the setting on both sides.

3.8 Leg rests



WARNING!

Risk of injury

 Never lift the wheelchair by the footrest supports or leg rests.



CAUTION!

Risk of pinching or crushing fingers

There are swing mechanisms where you could trap your fingers.

 Be careful when using, swinging, disassembling or adjusting these mechanisms.


ļ

NOTICE!

Risk of damage to the leg rest mechanism

 Do not place anything heavy, or let children sit on the leg rest.

3.8.1 Swing away leg rests



Swing outwards

1. Activate the release lever (A) and swing the leg rests to the outside.

Swing forwards

12

1. Swing the leg rest forwards until it engages.


Unhinging

- 1. Activate the release lever A.
- 2. Pull the leg rest upwards.

Hinging

1. Hinge the leg rest on the front of the frame and swing it forwards until it engages.

3.8.2 Swing away, Angle Adjustable Leg Rests



Swing outwards

1. Activate the release lever (A) and swing the leg rests outwards.

Swing forwards

1. Swing the leg rest forwards until it engages.

Unhinging

- 1. Activate the release lever (A), swing the leg rest outwards.
- 2. Pull the leg rest upwards.

Hinging

1. Push the leg rest down into its receiver and swing it forwards until it engages.

Adjusting angle




There are six preset positions available for angle adjustment.


- 1. Rotate the knob © with one hand while supporting the leg rest with your other hand.
- 2. When a suitable angle is obtained, let go off the knob and the leg rest will lock into the desired position.

Swinging calf pad and adjustment


• Swinging calf pad forwards (1) or backwards (2):




- 1. Swing calf pad © forwards (1).
- 2. Pull up the calf pad © and swing it backwards (2).
- Locking the calf pad swing backwards:



- 1. Extract the grey spacer open ring \otimes out of its location.
- 2. Reverse the grey spacer open ring.
- 3. Insert the grey spacer open ring \otimes as shown on the right side of the diagram.
 - Reverse this procedure to unlocking the calf pad swing backwards.
- Adjusting height of calf pad:



- 1. Loosen knob ©.
- 2. Adjust to the desired height and firmly tighten the knob.
- Adjusting depth of calf pad:



The calf pad has four depth adjustment options:

- 1. Remove the fixing screw (F) with 5 mm Allen key.
- 2. Adjust to one of the four positions and firmly tighten the fixing screw (F) (3–5 Nm).


Adjusting foot plate

There are three different foot plates available.


• Standard height-adjustable foot plates:



- 1. Loose the fixing screw B with 5 mm Allen key.
- 2. Adjust the height and let the screw catch one of the recesses on the foot plate tube.
- 3. Tighten the fixing screw B (3–5 Nm) in the desired position.
 - The distance between the lowest part of the footrest and the ground must be at least 50 mm.
- Depth-, Angle- (1) and Width- (2) adjustable foot plates:



- 1. Loosen the fixing screw G or H with 5 mm Allen key.
- 2. Adjust the depth and angle positions and firmly tighten the fixing screw G (12 Nm) or H (8–9 Nm) in the desired position.
- Width-adjustable foot plates (2):



- 1. Loosen the two fixing screws I with 5 mm Allen key.
- 2. Adjust the width position and firmly tighten the two fixing screws I (3–5 Nm) in the desired position.

Invacare recommends that foot plate adjustment is carried out by a qualified technician.

To ensure a good position of the feet, two types of straps can be provided; the heel strap (serial) and the calf strap (optional) attached to the leg rest support. Both adjustable by hook and loop fasteners or sliding buckle.

3.9 Anti-tipper

An anti-tipper prevents the wheelchair from tipping backwards.



WARNING! Risk of overturning

Anti-tippers that are incorrectly set or no longer working can lead to overturning.

- Always check that the anti-tipper is working before using the wheelchair and have it set or readjusted by a qualified technician when required.
- In some configurations, the static stability of the wheelchair may be lower than 10°; Invacare strongly recommends the use of anti-tipper (available as an option).



WARNING!

Risk of overturning


On uneven or soft ground, the anti-tip device can sink into potholes or directly into the ground, thereby curtailing or eliminating its safety function.

 Only use the anti-tipper when travelling on even and firm ground.

NOTICE!

Risk of damage

 Never step on the anti-tipper or use it as a tipping aid.



Activating the anti-tipper

- Release the spring button (A) and swivel the anti-tipper downwards until it engages and the front release button (B) snaps into the frame hole.
- 2. Ensure front release button ® protrudes fully through frame hole.



WARNING!

Risk of tipping

An activated anti-tipper can catch when negotiating a step or an edge.

 Always deactivate the anti-tipper before driving over a step or kerb.

Deactivating the anti-tipper

 Release the spring button (A) and swivel the anti-tipper upwards until it engages and the front release button (B) snaps into the frame hole.



WARNING!

Risk of overturning

- Always advise the user if you deactivate the anti-tippers.
- There is a red label on the housing of the anti-tipper, this label shows when the anti-tipper is inactivated. When the anti-tipper is correctly activated, the red label will be concealed.

Setting the height



WARNING!


Risk of overturning

Tipping angle can be reduced by decreasing the clearance of the anti-tip device (bottom from the wheels of anti-tip device to ground/floor).


- Have your anti-tip device set by a clinician/ qualified technician.
- In case of modification of your stability (new set up, new option, new add on,...), please contact your clinician/ qualified technician.
- This adjustment is required with reference to the position and diameter of the rear wheel as well as the user conditions and his particular safety limits.
- Make sure that the anti-tip device is fitted correctly on each side (if applicable), the spring button (a) and fixation (b) must be locked into position.

3.10 Seat

3.10.1 Adjusting the seat depth




- 1. Lift the seat (A) upwards in order to fold the chair slightly.
- 2. Release the front part of the seat ® from the hook and loop strips.



- 3. Adjust the depth by pulling the sides of the seat forwards or backwards in the rails of the side tubes ©. Make sure that both sides are evenly adjusted.
- 4. Unfold the chair again.
- 5. Attach the front part of the seat to the hook and loop strips again.

3.10.2 Adjusting the shape of the seat



- Release and pull back the upper section of the seat cover
- 2. Adjust the hook and loop strips ® to the desired tension.
- Fold the seat cover back in place.
 Always use a seat cushion when testing the final shape of the seat.



WARNING!

If the bands are set very loose, you could have pressure points on the thighs.

 Ensure that the bands are set correctly with your clinician.



NOTICE!

 Do not overtighten the bands; this is to ensure that the geometry of the wheelchair is not changed.

4 1608467-І



NOTICE!

 Only ever tighten the bands when the wheelchair is unfolded.

3.10.3 Seat cushion

A suitable cushion is recommended to provide an even pressure distribution on the seat.



CAUTION!

Risk of instability

Adding a cushion to the seat will raise your height above the ground and can affect your stability in all directions. If a cushion is changed it may also change the user's stability.

- If the thickness of the cushion is changed, a complete set up of the wheelchair need to be done by a qualified technician.
- We recommend using an Invacare or Matrx cushion with anti-slip base cover in order to prevent sliding.

4 Options

4.1 Comfort Version

It is made up of a rigid shell A and a preformed foam cushion B which is locked in place by hook and loop fasteners. The backrest shell is removable to allow the folding of the wheelchair. The comfort version is always delivered in passive rear wheel position for a better stability.



WARNING!

Risk of injury

Never lift the wheelchair by the backrest shell.




CAUTION!

Risk of pinching

 Keep your fingers away from movable parts during folding, disassembling or adjustment.

Remove the backrest shell



- 1. Slightly loosen the two hand screws ©.
- 2. Slide the levers
 inwards until they are in open position.
- 3. Disengage the hooks (E) from their supports.

Install the backrest shell

- 1. Engage the hooks (E) into their supports.
- 2. Slide the levers © outwards until they are in lock position.
- 3. Firmly tighten the two hand screws (A) and makes sure that the TWO hand screws (A) are properly tighten .


It is recommended that this operation be performed only by the attendant. It is important to ensure that the backrest shell is correctly fitted and engaged at all times when the chair is in use. Make sure the backrest cushion hook and loop fasteners are in good state and well positioned.

4.2 Transit version

The transit version is designed to be driven only by the attendant. To facilitate sideways transfers and save space, the wheelchair is equipped with rear wheels of 305 mm (12").

The distance between the brake shoe and the tyre can be adjusted. The adjustment must be carried out by a qualified technician.

The manual brakes (optional drum brake) are only accessible to the attendant; operate the handle (A) to lock the wheelchair in parking position.



- 1. 1. To engage the brake, pull the brake lever upwards as far as possible.
- 2. To disengage the brake, push the brake lever downwards.

For drum brake, please read 6.2 Braking during use, page 21.


Transit version equipped with a reclining backrest is always fitted with anti-tippers.




WARNING!

A loose belt can cause the user to slip down and create a risk of strangulation.

- The posture belt should be mounted by a qualified technician and fitted by a healthcare professional.
- Always make sure that the posture belt is tightly fitted across the lower pelvis.
- Each time the posture belt is used, check if it fits properly. Changing the seat and/or backrest angle, the cushion and even your clothes influence the fit of the belt.



Closing and opening the posture belt



Ensure that you are sitting fully back in the seat and that the pelvis is as upright and symmetrical as possible.


1608467-I 15

- 1. To close, push the catch (A) into the buckle clasp (B).
- 2. To open, push the PRESS button © and pull the catch (A) out of the buckle clasp (B).

Adjusting the length



The posture belt has a good length when there's just sufficient space for a flat hand between body and belt.




- 1. Shorten or extend the loop ${\Bbb D}$ as required.
- 2. Thread loop ① through catch ④ and plastic buckle ⑤ until the loop is flat.
- 3. Ensure loop ① is perfectly fitted in plastic buckle ⑤.
- 4. Secure adjustment with the strap extremity fitted in the buckle (E).



WARNING!


Risk of sliding and strangulation / Risk of falling

- Do the adjustments on both sides equally, so that the buckle clasp remains in a central position.
- Ensure that both strap extremities are threaded two times into the buckle E to avoid the belt from loosening.
- Make sure that the belts do not get caught in the spokes of a rear wheel.



4.3 Dual Handrim DHR

This quick release dual handrim system (left or right hand) allows the user to drive the wheelchair with a single arm using two handrims on the same wheel with 2 possible side positions of the small handrim. The user catches both handrims to propel the wheelchair forward or backwards; turns on the LH or RH side by gripping the large or the small handrim only. The DHR system is always delivered in passive rear wheel position for a better stability.





CAUTION! Risk of injury

 The change of control side and position of the small handrim must be carried out by a qualified technician.



WARNING! Risk of injury

Always remove the telescopic rear wheel shaft
 © before removing the rear wheels.



WARNING! Risk of injury

The wheels come off

 Check that the rear wheel is securely locked in position! It should not be possible to remove the wheels when the removable axle button A is inactivated. Test this by trying to pull the wheel off



CAUTION! Risk of injury

Minor bruises, pain

 Be careful not to get your fingers trapped between the rear wheel spokes and the three brackets of the outer handrim.

Disengaging and Engaging the Rear Wheel Shaft

To fold the wheelchair it is necessary to disengage the telescopic rear wheel shaft ©.

- Disengage the telescopic shaft © by sliding one part inside the other
- 2. To facilitate this operation move the small handrim forwards to backwards.
- Reverse procedure to engage the telescopic rear wheel shaft.




CAUTION!

Risk of accident

 The telescopic rear wheel shaft © is an integral part of the wheelchair and the user will be unable to propel the wheelchair without it.

Removing and Installing the Rear Wheel



The removable and interchangeable rear wheels facilitate transport as well as change of the control side.

- 1. Remove the telescopic rear wheel shaft ©.
- 2. Press the removable axle button **(A)**. Keep it pressed and pull the wheel outwards.
- 3. Remove the rear wheel.
- 4. Repeat procedure on the opposite side.
- 5. Reverse procedure to reinstall the rear wheels.

16 1608467-I

4.4 One arm drive by lever

The One arm drive by lever control system (left or right hand) allows the user with low muscular tone to propel the wheelchair with a single arm. The user grips the drive motion lever which propels the wheelchair frontwards or backwards (reversing gear lever (A), central position is the neutral gear) thanks to an in and out movement; steering is performed by simply turning the drive motion lever to the LH/RH side and braking by pulling the drive motion lever backwards. The OADL system is always delivered in passive rear wheel position for a better stability.

- The drive motion lever can be removed to facilitate transfer or to move next to a table; loosen the handle lever (B) and pull up the drive motion lever. Reverse the procedure to reinstall.



CAUTION!

Risk of pinching or crushing

There may be a very small gap between the rear wheel and the transmissions links with the risk that you could trap your fingers.

 Keep your fingers away from movable parts when using the OADL system, always keep your hand on the hand lever.

Adjust the propulsion effort



The adjustment is performed by moving the transmission links sleeve on the drive motion lever.

- 1. Slightly loosen the handle lever ©.
- Slide the transmission links sleeve downwards to reduce the stress.
- 3. Firmly tighten the handle lever © and makes sure that the handle lever © is properly tighten.



CAUTION!

Risk of pinching or crushing

Make sure that the 2 handle levers ® and © are properly positioned so that they are not bothersome or hurtful for the user and attendant.

Disengage the steering mechanism

It may be necessary, for the attendant to move the wheelchair, to disengage the ball joint system from the fork in order to release the front directional wheel.

- 1. Press the end of the disengaging spring ①.
- 2. Release the vertical axle (E) from its support.
- 3. Reverse the procedure to reinstall.

4.5 Posture belt

The wheelchair can be equipped with a posture belt. It prevents the user from sliding downward in the wheelchair or from falling out of the wheelchair. The posture belt is not a positioning device.


\triangle

WARNING!

Risk of serious injury / strangulation

A loose belt can allow the user to slip down and create a risk of strangulation.

- The posture belt should be mounted by a qualified technician and fitted by the responsible prescriber.
- Always make sure that the posture belt is tightly fitted across the lower pelvis.
- Each time the posture belt is used, check if it fits properly. Changing the seat and/or backrest angle, the cushion and even your clothes influence the fit of the belt.






WARNING!

Risk of serious injury during transport

In a vehicle, a user in his wheelchair must be secured by a safety belt (3-point belt). A posture belt only is not sufficient as a personal restraint device.

 Use the posture belt as a complement, but not as a substitute to the 3-point safety belt, when transporting the wheelchair user in a vehicle.

Closing and opening the posture belt




Ensure that you are sitting fully back in the seat and that the pelvis is as upright and symmetrical as possible.

- 1. To close, push the catch (A) into the buckle clasp (B).
- 2. To open, push the PRESS button \bigcirc and pull the catch \bigcirc out of the buckle clasp \bigcirc .


Adjusting the length

]

The posture belt has good length, when there's just sufficient space for a flat hand between body and belt.



- 2. Thread loop @ through catch $ext{\@A}$ and plastic buckle $ext{\@E}$ until the loop is flat.
- 3. Ensure loop ① is perfectly fitted in plastic buckle ⑤.
- 4. Secure adjustment with the strap extremity fitted in the buckle (E).




Fitting the posture belt at the fixations



CAUTION!

Risk of serious injury during transport


- Thread the belt loop through both plastic buckles to avoid the belt from loosening.
- Do the adjustments on both sides equally, so that the buckle clasp remains in a central position.
- Ensure strap extremities are fitted in the buckle
- Make sure that the belts do not get caught in the spokes of a rear wheel.



4.6 Headrest


The height and the removal are operated by the handwheel. The bar is equipped with an adjustable stop block.

Height adjustment



- 1. Loosen the screw in the stop block **B**.
- 2. Loosen the handwheel A.
- 3. Adjust the headrest to the desired position.
- 4. Re-tighten the handwheel (A) down to the top of the headrest attachment.
- 5. Slide the stop block [®].
- 6. Re-tighten the screw.
 - It is now possible to remove the headrest and reinsert it in the desired position without further adjustments.

Depth / angle adjustment



- 1. Loosen the handle D.
- 2. Loosen the screws ©.
- 3. Adjust the depth and angle of the headrest.
- 4. Re-tighten the handle and the screws.

4.7 Lateral positioning pads (for adjustable backrest cover only)



WARNING!

Risk of trapping

 Be careful not to trap your trunk between the two lateral positioning pads during front or lateral transfers; when the user is moving in the wheelchair.




CAUTION!


Risk of uncomfortable posture

Insufficient space between the lateral positioning pads can be uncomfortable for certain users.

- This adjustment must be performed by a qualified technician upon agreement by a prescribing physician.
- Consult with your provider.
- The lateral positioning pads can be adjusted in height and sideways.



- 1. Remove the backrest cushion A.
- 2. Adjust the lateral positioning pads (B) in height and sideways to the required position with the hook surface of the backrest cover (C).
- 3. Ensure that the two hook surfaces (pads and back cover) are set correctly.
- 4. Put the backrest cushion A back in place.



- The lateral positioning pads \circledR must be located above the arm pad ข in order to be able to fold the wheelchair.
- If the seat width is larger then 430 mm (E), the lateral positioning pads (B) must be removed before folding the wheelchair.

4.8 Tipper aid


With the tipper aid, an assistant can tip the wheelchair more easily, in order to navigate steps for example.



WARNING!

Risk of overturning

 Ensure that the tipper aid does not project beyond the external diameter of the rear wheel.



- 1. Hold the wheelchair by the push handles.
- 2. Press the tipper aid with your foot and hold the wheelchair in the tipped position until you have overcome the obstacle.

4.9 Table tray



WARNING!

Risk of tipping / injury

Max load on the table: 8 kg




CAUTION!

Risk of discomfort / minor bruises

 Make sure that the elbows of the user are placed on the table when pushing the wheelchair. If the elbows protrudes from the table while pushing the wheelchair, there is a risk of discomfort or minor bruises.

Depth adjustment



Slide the table tray forwards or backwards to adjust to the required depth.



CAUTION!

Risk of discomfort

 When adjusting the depth of the table, make sure not to squeeze the stomach of the user.

4.10 Half tray



WARNING!

Risk of falling / injury

 The half tray must never be used as a replacement for the posture belt.



WARNING!

Risk of tipping / injury

Max load on the half tray: 1.5 kg



CAUTION!

Risk of discomfort / minor bruises

 Make sure that the elbow of the user are placed on the half tray when propelling the wheelchair.
 If the elbow protrudes from the wheelchair while propelling the wheelchair, there is a risk of discomfort or minor bruises.



CAUTION!

Risk of pinching fingers


There is a mechanism below the half tray where you could trap your fingers.

- Be careful when adjusting the tray or when using the swing-away function.
- The half tray can be adjusted in width and swung away.

Mounting the half tray




1. Insert the tube of the half tray (A) into its support (B) below the arm pad.



- 2. Push on the push pin c and insert the ring D on the tube A.
- 3. Release the push pin ©.

Adjusting the depth of the half tray



1. Slide the half tray (A) forwards or backwards to adjust to the required depth.



CAUTION!

Risk of discomfort


 When adjusting the depth of the half tray, make sure not to squeeze the stomach of the user.

Swing away the half tray



4.11 Add a half tray cushion

A cushion for the half tray can be placed on the half tray in order to have a softer surface for the arms and elbows.



- 1. Place the cushion for the half tray (A) on the tray.
- 2. Place the attachment straps 8 into the plastic buckle and around the half tray A.

4.12 Drip stand




WARNING!

Risk of tipping / injury

Max load on the drip stand: 4 kg (2 x 2 kg)

The rod of the drip stand must always be placed in a vertical position, i.e in a 90 degree angle to the ground, no matter the position of the backrest or the wheelchair.



Adjusting the Height and Angle



- 1. Loosen the lever A.
- 2. Adjust the drip stand to the desired height or/and angle.
- 3. Firmly tighten the lever (A).

4.13 Transfer kit wheels

If your wheelchair is too wide for certain applications, e.g. trains/aeroplanes, narrow passageways, narrow doors, the transfer kit wheels can be used.



WARNING!

 When using transfer kit wheels, the parking brakes no longer have any effect and you cannot control your wheelchair via the handrims.

The transfer kit wheels are attached directly to the rear frame as an option.



WARNING!

Risk of tipping sidewards

 Make sure that anti-tippers with transfer kit wheels are installed on both sides of the wheelchair.

Switching to Transfer kit wheels

- 1. Activate the anti-tippers with transfer kit wheels on both sides, $\rightarrow 3.9$ Anti-tipper, page 13.
- 2. With the help of an assistant, remove the rear wheel,
 → 7.4 Removing and fitting the rear wheels, page 25 and lower the wheelchair to the transfer kit wheel.
- 3. Repeat the process on the other side.

Switching from Transfer kit wheels to rear wheels

- With the help of an assistant, fit the rear wheel back onto the removable axle, → 7.4 Removing and fitting the rear wheels, page 25, and lower the wheelchair to the rear wheel.
- 2. Repeat the process on the other side.

4.14 Air pump

The air pump is fitted with a universal valve connection. Using the correct valve connection depends of the inner tube valve type.

- 1. Remove the dust cap from the valve connection.
- 2. Push the valve connection onto the open valve on the wheel and pump the wheel up.
 - There are two types of pumps available:
 - Low pressure <6 bars
 - High pressure >6 bars

Instruction for use printed on the packaging.

20 1608467-І

4.15 Passive illumination

You can attach reflectors to the rear wheels.

4.16 Cane holder





WARNING!

Risk of overturning

- Ensure that the cane holder does not project beyond the external diameter of the rear wheel.
- 1. Place the cane in the holder B.
- 2. Fix the top of the cane to the backrest **(A)**.

5 Setup

5.1 Safety Information



CAUTION!

Risk of injury

 Before using the wheelchair, check its general condition and its main functions in chapter 8.2 Maintenance schedule, page 28.



WARNING!

- Check all parts for shipping damage before use.
- In case of damage, do not use the device.
 Contact your Invacare provider for further instructions.

Your authorized provider will supply your wheelchair ready for use and will explain the main functions, ensuring that the wheelchair meets your needs and requirements and it is properly adjusted for safe use.

If you receive your wheelchair folded, read 7.2 Folding and unfolding the wheelchair, page 25.

5.2 Delivery Check

Any transport damage must be reported immediately to the transport company. Remember to keep the packaging until the transport company has checked the goods and a settlement has been reached.

6 Using the wheelchair

6.1 Safety Information



WARNING!

Risk of accidents

Uneven tyre pressure can have a huge effect on handling.

Check the tyre pressure before each journey.



WARNING!

Risk of falling out of the wheelchair

When using front wheels that are too small, the wheelchair could get stuck at curbs or in floor grooves.

 Make sure that the front wheels are suitable for the surface you're driving on.



CAUTION!

Risk of crushing

There may be a very small gap between the rear wheel and the mudguard or armpad with the risk that you could trap your fingers.

Ensure that you always propel your wheelchair using the handrims only.

6.2 Braking during use

Whilst you are moving, you brake by transferring force to the handrim with your hands.



WARNING!

Risk of accidents

If you engage the parking brakes while you are moving, the direction of movement can become uncontrollable and the wheelchair may stop suddenly, which can lead to a collision or to you falling out.

 Never engage the parking brakes while you are moving.



WARNING!

Risk of falling of the wheelchair

If the wheelchair is rapidly decelerated by an assistant pulling at the push handles, the user may fall out of the wheelchair.

- Always apply the posture belt if present
- Make sure your assistant is qualified in transferring occupied wheelchairs.
- Operate the two brakes simultaneously and do not take slopes exceeding 7° to ensure perfect control of the wheelchair steering.
- Always use the parking brakes to secure the wheelchair during short or long stops.



CAUTION!


Risk of burning your hands

If you brake for a long time, a lot of frictional heat is produced at the handrims (especially anti slip handrims)

- Wear suitable gloves.
- 1. Hold the handrims and press evenly with both hands until the wheelchair stops.

Braking with an assistant

With an assistant's brake (drum brake) an assistant can brake the wheelchair while it is moving. The assistant's brake can also be used as a parking brake.



- 1. To brake, pull the brake lever (A) on the push handle.
- 2. To park, pull the brake lever firmly and move the safety lever (B) upwards.
- To release, pull the brake lever until the safety lever disengages.

6.3 Move to/from the wheelchair



WARNING!

Risk of burning your hands

There is a high risk of overturning during the transfer.

- Only get in and out without assistance if you are physically able to do so.
- When transferring, position yourself as far back as possible in the seat. This will prevent damaged upholstery and the possibility of the wheelchair tipping forward.
- Make sure that both castors are facing straight to the front.



WARNING!

Risk of overturning

The wheelchair could tip forwards if you stand on the footrest.

Never stand on the footrest when getting in and



CAUTION!

If you disengage or damage the brakes the wheelchair could roll away out of control.

 Do not support yourself on the brakes when getting in and out.



NOTICE!

The mudguards and armrests could become damaged.

 Never sit on the mudguards or armrests when getting in and out.



- 1. Propel the wheelchair as near as possible to the seat that you want to move to.
- 2. Engage the parking brakes.

- Remove the armrests or move them upwards out of the way.
- 4. Detach the leg rests or swing them outwards.
- 5. Place your feet on the ground.
- 6. Hold the wheelchair and, if necessary, also hold a fixed object in the vicinity.
- 7. Move slowly to chair.

6.4 Driving and Steering the Wheelchair

You drive and steer the wheelchair using the handrims. Before driving without an assistant you must find your wheelchair's tipping point.



WARNING!

Risk of tipping

The wheelchair can tip backwards.

- When finding the tipping point, an assistant must stand immediately behind the wheelchair to catch it if it tips over.
- To prevent tipping, install an antitipper device.



WARNING!

Risk of tipping

The wheelchair can tip forward.

 When setting up your wheelchair, test its behaviour in terms of tipping forward and adjust your driving style accordingly.



CAUTION! Risk of tipping

A heavy load hanging on the backrest can affect the wheelchair's centre of gravity.

Change your driving style accordingly.

Finding the Tipping Point



- 1. Release the brake.
- 2. Roll backwards a short distance, grasp both handrims firmly and push forwards with a slight kick.
- 3. The shift in weight and steering in opposite directions with the handrims will enable you to identify the tipping point.

6.5 Negociating steps and kerbs



WARNING!

Risk of overturning

When negotiating steps you could lose your balance and tip the wheelchair over.

- Always approach steps and kerbs slowly and carefully.
- Do not go up or down steps that are higher than 25 cm.

22 1608467-1



CAUTION!

An activated anti-tip device prevents the wheelchair from tipping backwards.

 Deactivate the anti-tip device before going up or down steps or kerbs.



WARNING!

Risk of injury to the assistant and damage to the wheelchair

Tilting the chair with a heavy user can hurt the assistant's back and damage the chair.

 Make sure to be able to safely control the wheelchair with a heavy user before performing a tilting manoeuvre.

Going down a step with an assistant



- Move the wheelchair right up to the kerb and hold the handrims.
- 2. The assistant should hold both push handles, place one foot on the tipper aid (if installed) and tilt the wheelchair backwards so that the front wheels lift off the ground.
- The assistant should then hold the wheelchair in this position, push it carefully down the step and then tilt it forwards until the front wheels are back on the ground.

Going up a step with an assistant



WARNING!

Risk of serious injuries

Going up steps and kerbs frequently can result in earlier than expected fatigue break of the wheelchair backrest. The user might fall out of the wheelchair..

- Always use a tipper aid when going up steps or kerbs.
- 1. Move the wheelchair backwards until the rear wheels touch the kerb.
- 2. The assistant should tilt the wheelchair using both push handles so that the front wheels lift off the ground, then pull the rear wheels over the kerb until the front wheels can be placed back on the ground.

Going down a step without an assistant





WARNING!

Risk of tipping

When going down a step without an assistant you could tip over if you cannot control your wheelchair.

- First learn how to go down a step with an assistant.
- Learn how to balance on the rear wheels, 6.4
 Driving and Steering the Wheelchair, page 22.
- 1. Move the wheelchair right to the kerb, lift the front wheels and keep the wheelchair balanced.
- 2. Now slowly roll both rear wheels over the kerb. While doing this, hold the handrims firmly with both hands until the front wheels are back on the ground.

6.6 Going up and down stairs



WARNING!

Risk of overturning

When negotiating stairs you could lose your balance and overturn your wheelchair.

 Always use two assistants when negotiating stairs with more than one step.



 You can go up and down stairs by taking them one step at a time, as described above. The first assistant should stand behind the wheelchair holding the push handles. The second assistant should hold a solid part of the front frame to steady the wheelchair from the front.

6.7 Negotiating Ramps and Slopes



WARNING!

Risk due to wheelchair being out of control When negotiating slopes or gradients your wheelchair could tip backwards, forwards or sideways.

- Always have an assistant behind the wheelchair when approaching long slopes.
- Avoid lateral slopes.
- Avoid slopes of more than 7°.
- Avoid jerking when changing direction on a slope.



CAUTION!

Your wheelchair could run away even on slightly sloping ground if you do not control it using the handrims.

 Use the parking brakes if your wheelchair is stationary on sloping ground.

Going up slopes

To go up a slope, you must create some momentum, keep up the momentum and control the direction at the same time.



1. Bend your upper body forwards and propel the wheelchair with quick, powerful strokes on both handrims.

Going down slopes

When going down slopes, it is important to control your direction and particularly your speed.



1. Lean back and carefully allow the handrims to run through your hands. You should be able to stop the wheelchair at any time by gripping the handrims.



CAUTION!

Risk of burning your hands

If you brake for a long time, a lot of frictional heat is produced at the handrims (especially anti slip handrims).

Wear suitable gloves.

6.8 Stability and balance when seated

Some everyday activities and actions require you to lean forwards, sideways or backwards out of the wheelchair. This has a major effect on the wheelchair's stability. To keep your balance at all times, proceed as follows:

Leaning forwards



WARNING! Risk of falling

If you lean forwards out of the wheelchair you could

- Never bend too far forwards and do not shift forwards in your seat to reach an object.
- Do not bend forwards between your knees to pick up something off the floor



- 1. Point the front wheels forwards. (To do this, move your wheelchair forwards slightly then back again.)
- 2. Apply both parking brakes.
- 3. When leaning forwards your upper body must remain over the front wheels.

Reaching backwards



WARNING! Risk of falling

If you lean too far backwards you could tip your wheelchair over.

- Do not lean out over the backrest.
- Use an anti-tip device.



- 1. Point the front wheels forwards. (To do this, move your wheelchair forwards slightly then back again.)
- 2. Do not apply the parking brakes.
- 3. When reaching backwards do not reach so far that you have to change your sitting position.

7 Transport

7.1 Safety Information



WARNING!

Risk of injury if the wheelchair is not properly secured

In the event of an accident, braking manoeuvre, etc. you may suffer serious injuries from flying wheelchair parts.

 Firmly secure all wheelchair components in the means of transport to prevent them from coming loose during the journey.

NOTICE!

Risk of injury if the wheelchair is not properly secured

Excessive wear and abrasion could affect the strength of load-bearing parts.

 Do not pull your wheelchair across abrasive surfaces without the wheels fitted (e.g. pulling the frame over tarmac).

1608467-I

7.2 Folding and unfolding the wheelchair

Folding the wheelchair



- 1. Remove the seat cushion, if present.
- 2. Fold the backbrace, if present.
- 3. Fold the foot plates upwards.
- 4. Pull the seat cover upwards.
- 5. You can fold the backrest frontwards, if your wheelchair is fitted with an angle adjustable and fold down backrest.
- 6. Pull the pull-cord upwards at the back of the wheelchair. The angle adjustable and fold down backrest folds forwards.

Unfolding the wheelchair





CAUTION!

Risk of getting fingers caught

There is a risk of getting your fingers caught between the seat edge and the frame.

- Never clasp the seat edge with your fingers.
- 1. Position the wheelchair next to you.
- 2. Tilt the wheelchair, if possible, gently towards you.
- 3. With your hand flat, press the seat edge closest to you, so that the seat surface flattens out.
- 4. Put the wheelchair back completely on the floor and check that the seat edges on both sides are in the guide components.
- 5. If your wheelchair is fitted with an angle adjustable and fold down backrest, pull up the backrest at the push-handles until the ratchet bolts engage in the lateral supports.
- 6. Fold the foot plates down.
- 7. Unfolding the backbrace, if present.

7.3 Lifting the wheelchair



NOTICE!

Never lift the wheelchair by removable parts (armrests, footrests).

Ensure the backrest posts are securely in place.



- 1. Fold the wheelchair, see 7.2 Folding and unfolding the wheelchair, page 25.
- Always lift the wheelchair by gripping the frame at pointsA.

7.4 Removing and fitting the rear wheels




WARNING!

Risk of overturning

If the removable axle of a rear wheel is not fully engaged, the wheel can become loose during use. This can lead to overturning.

 Always ensure that the removable axles are fully engaged whenever you fit a wheel.



Removing the rear wheels

- 1. Release the brakes.
- 2. With one hand, hold the wheelchair upright.
- 3. With the other, hold the wheel through the external spoke rim around the wheel hub.
- Using your thumb, press the removable axle button (A).
 Keep it pressed and pull the wheel out of the adapter sleeve (B).

Fitting the rear wheels

- 1. Release the brakes.
- 2. With one hand, hold the wheelchair upright.
- 3. With the other, hold the wheel through the external spoke rim around the wheel hub.
- 4. Using your thumb, press the removable axle button and hold it down.
- 5. Push the axle into the adapter sleeve (B) up to the stop.
- 6. Release the removable axle button and make sure that the wheel is secure.

7.5 Transporting the Manual Wheelchair without Occupant



CAUTION! Risk of injury

 If you are unable to fasten your manual wheelchair securely in a transport vehicle, Invacare recommends that you do not transport it.

Your manual wheelchair may be transported without restrictions, whether by road, rail or by air. Individual transport companies have, however, guidelines which can possibly restrict or forbid certain transport procedures. Please ask the transport company regarding each individual case.

• Invacare strongly recommends securing the manual wheelchair to the floor of the transporting vehicle.

7.6 Transporting the Occupied Wheelchair in a Vehicle

Even when the wheelchair is properly secured and the following rules are met, injuries to passengers may occur in a collision or sudden stop. Therefore Invacare strongly recommends transferring the wheelchair user to the vehicle seat with the 3-point passenger restraint system on. Do not make alterations or substitutions to points of the wheelchair (structure, frame or parts) without the written consent of Invacare Corporation. The wheelchair has been successfully tested according to the requirements of ISO 7176–19 (Frontal impact).



WARNING!

Risk of serious injury or death

To use the wheelchair as a seat in a vehicle the backrest height minimum must be at least 400 mm.

To transport the wheelchair with user in a vehicle, a restraint system must be installed in the vehicle. Wheelchair tie-downs and occupant restraint systems must be approved according to ISO 10542-1. Contact your local Invacare authorized provider for more information on getting and installing an approved and compatible restraint system.



WARNING!

If, for some reason, it is impossible to transfer the wheelchair user to a vehicle seat, the wheelchair can be used as a seat in a vehicle if the following procedures and regulations are followed. A transportation kit (optional) have to be fitted on the chair for such a purpose.

- The wheelchair must be secured in the vehicle with a 4-point wheelchair restraint system.
- The user must wear a 3-point passenger restraint system secured to the vehicle.
- The user can be additionally secured with a posture belt in the wheelchair.



WARNING!

Safety restraint devices must only be used when the wheelchair user's weight is 22 kg or more (ISO-7176-19).

 Do not use the wheelchair as a seat in a vehicle when the user weight is lower than 22 kg.



WARNING!

- Before journey contact transporter and request information about the capability for the below required equipment.
- Make sure there is sufficient free space around the wheelchair and user to avoid the user making contact with other vehicle occupants, unpadded parts of the vehicle, wheelchair options or anchor points of the restraint system.



WARNING!

- Make sure the tie-down points at the wheelchair are not damaged and that the parking brakes are fully functional.
- It's recommended to use puncture-proof tyres during transport to avoid brake problems due to reduced tyre pressure.



WARNING!

Injury or damage may occur from wheelchair components or options loosened during a collision or sudden stop.

- Ensure all removable or detachable components and options are removed from the wheelchair and securely stored in the vehicle.
- It is essential to have your wheelchair checked by a qualified technician after an accident, collision etc.

Ramps and Slopes:



WARNING! Risk of injury

The wheelchair can move forward / backward uncontrollably by mistake.

- Never leave the user unattended when transporting the wheelchair up or down ramps or slopes.
- Therefore Invacare strongly recommends transferring the wheelchair user to the vehicle with the posture belt on.



NOTICE!

- Refer to the user manuals supplied with the restraint systems.
- The following illustrations may differ depending on the restraint system supplier.
- The choice of wheelchair configuration (seat width and depth, wheelbase) influences maneuverability and access to motor vehicles.

Securing the wheelchair with a 4-point restraint system



26 1608467-I




WARNING!

- Place the wheelchair with the user forwardfacing in direction of vehicle travel.
- Engage the wheelchair parking brakes.
- Activate anti-tipper device (if installed).

The wheelchair tie-down positions where the restraint system straps must be placed are marked with snap hook symbols (see following figures in Labels and symbols on the product section).


1. Using the front and rear straps of the 4-point restraint system, secure the wheelchair to the vehicle mounted rails. Refer to the user manual supplied with 4-point restraint system.

Front side tie-down positions for belt straps:



- 1. Attach the front straps above the castor supporters as shown in the figure above (see location of the attachment labels).
- 2. Attach front straps to the rail system referring to best practice recommended instructions from the 4-point restraint system manufacturer.
- 3. Disengage parking brakes and apply tension front straps by pulling the wheelchair backwards from the rear.
- 4. Re-engage parking brakes.

Rear side tie-down positions for snap hooks:



- 1. Attach the snap hooks to the orange rings as shown in the two figures above (see location of the attachment labels).
- 2. Attach rear straps to the rail system referring to best practice recommended instructions from the 4-point restraint system manufacturer.
- 3. Tighten the straps.



NOTICE!

 Make sure that the snap hooks are covered with slip resistant material to avoid laterally slipping on the axle.



NOTICE!

- Check the plungers are fully engaged on both sides and located in the same position of the cut-out section of the rail.
- Make sure that the angle between the rails and the straps lies between 40° and 45°.

Fit posture belt



WARNING!

- The posture belt can be used in addition to but never as a substitute for an approved 3-point passenger restraint system.
- 1. Adjust posture belt to fit the wheelchair occupant, see 4.5 *Posture belt, page 17*.

Fit specific attachment kit for Comfort version only



WARNING!

A specific attachment kit (available as an option) is mandatory to be used for car transportation, no other attachment is permitted! A mounting instruction is delivered with the kit, contact your authorized provider for further information.

Fastening the 3-point passenger restraint system





NOTICE!

 The illustration above may differ depending on the restraint system supplier.



WARNING!

- Ensure the 3-point passenger restraint system fits as tightly across the user's body as possible without discomfort and no part is twisted.
- Ensure the 3-point passenger restraint system is not held away from the user's body by parts of the wheelchair such as armrests or wheels etc.
- Ensure the 3-point passenger restraint system has a clear path from the user to the anchor point without interference by any part of the vehicle, wheelchair, seating or option.
- Ensure the pelvic belt fits snugly over the occupant's pelvis and is not allowed to ride up into the abdominal area.
- Ensure the user is able to reach the release mechanism unaided.





1. The 3-point belt restraint should fit as tightly across the user's body as possible without discomfort.

NOTICE!

 Ensure that the belt webbing is not twisted when in use.



NOTICE!

 Apply the pelvic belt of the 3-point passenger restraint system low across the pelvis so that the angle of the pelvic belt is within the preferred zone (a) of 30° to 75° to the horizontal. A steeper angle is preferred, but never exceeding 75°.



- The image shows an incorrect placement of three-point belt restraint.
- 2. The belt restraint must not be held away from the user's body by parts of the wheelchair such as armrests or wheels etc.
- 3. If the vehicle is equipped with a head restraint and the wheelchair with the head/neck support, then assess with a clinician which one should be used.
 - The Invacare head and neck supports are not restraint devices. Nevertheless, they have been crash tested to verify their attachment points.

8 Maintenance

8.1 Safety Information



WARNING!

Some materials deteriorate naturally over time.

This could result in damage to wheelchair components.

 Your wheelchair should be checked by a qualified technician at least once a year or if it has not been used for a long period.

Packaging for return

The wheelchair shall be sent to the authorized provider in an appropriate packaging to avoid damage during the shipment.

8.2 Maintenance schedule

To ensure safe and reliable operation, carry out the following visual checks and maintenance regularly or have it carried out by another person.

	weekly	monthly	annually
Check the tyre pressure	х		
Check that rear wheels are seated correctly	х		
Check posture belt	х		
Check the folding mechanism		х	
Check the castors and its fixation		х	
Check bolts		х	
Check spokes		х	
Check parking brakes (mechanism and cable)		x	
Have wheelchair checked by a qualified technician			х

General visual check

- Examine your wheelchair for loose parts, cracks or other defects.
- 2. If you find anything, stop using your wheelchair immediately and contact an authorized provider.

Check the tyre pressure

- 1. Check the tyre pressure, see section "Tyres" for more information.
- 2. Inflate the tyres to the required pressure.
- 3. Check the tyre tread at the same time.
- 4. If necessary, change the tyres.

Check that rear wheels are seated correctly

- Ensure that the axle ball is outside its receiver by visually checking it and pull on the rear wheel to check that the removable axle is seated correctly. The wheel should not come off.
- 2. If the rear wheels are not engaged properly, remove any dirt or deposits. If the problem persists, have the removable axles re-fitted by a qualified technician.

Check posture belt

ļ

NOTICE!


- Loose posture belts must be adjusted by an authorized provider.
- Damaged posture belts must be replaced by a qualified technician.

Check the folding mechanism

1. Check that the folding mechanism is easy operating.

Check the castors and its fixation

- 1. Check that the castors turn freely.
- 2. Push and pull in the 3 directions (front and rear; left and right; up and down) to observe that is no free play and movement, also check that there is no visual damage.



- 3. Remove any dirt or hair from the castor bearings.
- 4. Faulty or worn out castors fixation must be replaced by a qualified technician.

Check the bolts

Bolts can work loose through constant use.

- 1. Check that the bolts are tight securely and without free play (on the footrest, castor, castor housing, seat cover, sides, backrest, frame, seat module).
- 2. Tighten any loose bolts with the suitable torque.

ļ

NOTICE!

Self-locking screws/nuts or thread-locking adhesive are used for several connections. If these are loosened, they must be replaced by new self-locking screws/nuts or they must be secured using new thread-locking adhesive.

Self-locking screws/nuts must be replaced by a qualified technician

Check the spoke tension

- 1. The spokes should not be loose or distorted.
- 2. Broken spokes must be replaced by a qualified technician.
- 3. Loose spokes must be tightened by a qualified technician.

Check the parking brakes

- 1. Check that the parking brakes are positioned correctly. The brake is set correctly if the brake shoe depresses the tyre by a few millimetres when the brake is engaged.
- 2. If you find that the setting is not correct, have the brakes correctly adjusted by a qualified technician.



NOTICE!

The parking brakes must be reset after replacing the rear wheels or changing their position.

Checking after a heavy collision or blow



NOTICE!

The wheelchair can sustain visibly undetectable damage as a result of a heavy collision or hard blow.

 It is essential to have your wheelchair checked by a qualified technician after a heavy collision or hard blow.

Repairing or changing an inner tube

- Remove the rear wheel and release any air from the inner tube.
- 2. Lift one tyre wall away from the rim using a bicycle tyre lever. Do not use sharp objects such as a screwdriver which could damage the inner tube.
- 3. Pull the inner tube out of the tyre.
- 4. Repair the inner tube using a bicycle repair kit or, if necessary, replace the tube.
- 5. Inflate the tube slightly until it becomes round.
- 6. Insert the valve into the valve hole on the rim and place the tube inside the tyre (the tube should lie right round the tyre with no creases).
- 7. Lift the tyre wall over the edge of the rim. Start close to the valve and use a bicycle tyre lever. When doing this, check all the way round to ensure that the inner tube is not trapped between the tyre and the rim.
- 8. Inflate the tyre to the maximum operating pressure. Check that no air is escaping from the tyre.



Spare parts

All spare parts may be obtained from an Invacare authorized provider.

8.3 Cleaning and disinfection

8.3.1 General Safety Information



CAUTION!

Risk of Contamination

Take precautions for yourself and use appropriate protective equipment.



NOTICE!

Wrong fluids or methods can harm or damage the product.

All cleaning agents and disinfectants used must be effective, compatible with one another and must protect the materials they are used to clean.

- Never use corrosive fluids (alkalines, acid etc.) or abrasive cleaning agents. We recommend an ordinary household cleaning agent such as dishwashing liquid, if not specified otherwise in the cleaning instructions.
- Never use a solvent (cellulose thinner, acetone etc.) that changes the structure of the plastic or dissolves the attached labels.
- Always make sure that the product is completely dried before taking it into use again.



For cleaning and disinfection in clinical or long-term care environments, follow your in-house procedures.

8.3.2 Cleaning Intervals

NOTICE!

Regular cleaning and disinfection enhance smooth operation, increases the service life and prevents contamination.

- Clean and disinfect the product:
- Regularly while in use,
- Before and after any service procedure,
- When it has been in contact with any body fluids.
- Before using it for a new user.

8.3.3 Cleaning

NOTICE!

Dirt, sand and seawater can damage the bearings and steel parts can rust if the surface is damaged.

- Only expose the wheelchair to sand and sea water for short periods and clean it after every trip to the beach.
- If the wheelchair is dirty, wipe off the dirt as soon as possible with a damp cloth and dry it carefully.
- Remove any options fitted (only options which do not require tools).
- Wipe down the individual parts using a cloth or soft brush, ordinary household cleaning agents (pH = 6 - 8) and warm water.
- 3. Rinse the parts with warm water
- 4. Thoroughly dry the parts with a dry cloth.
 - Car polish and soft wax can be used on painted metal surfaces to remove abrasions and restore gloss.

Cleaning upholstery

For cleaning upholstery refer to the instructions on the labels of the seat, cushion and backrest cover.

If possible, always overlap hook and loop strips (the self-gripping parts) when washing, to minimize lint and thread build-up on hook strips and prevent damage to upholstery fabric by these.

8.3.4 Washing

- Remove all loose and removable covers and wash them in a washing machine according to the washing instructions for each cover.
- Remove all padded parts such as seat cushions, armrests, headrest or neckrest with fixed padded parts, calf pads and so on and clean them separately. The padded parts can not be cleaned with a high-pressure cleaner or water jet.
- 3. Spray the wheelchair chassis with detergent, for example a car-cleaning agent with wax, and leave on to work.
- 4. Rinse the wheelchair chassis with a high-pressure cleaning or ordinary jet of water depending on how dirty the wheelchair is. Do not aim the jet towards bearings and draining holes. If the wheelchair chassis is washed in a machine the water must not be hotter than 60° C.
 - n Only use water and soft soap to clean the table.

 Leave the wheelchair to dry in a drying cabinet. Remove parts where water has collected for example in end tubes, ferrules etc. If the wheelchair has been washed in a machine, blow-drying with compressed air is recommended.

Multi stretch polyurethane (PU) coated fabric

Lighter stains on the fabric may be neutralized with a soft damp cloth and some neutral detergent. To neutralize larger, more persistent stains, wipe the fabric with alcohol or turpentine substitutes, and wash with hot water and a neutral detergent.

The fabric can be washed at temperatures up to 60° C. Normal detergents can be used.

All parts of the wheelchair with multi stretch polyurethane (PU) coated fabric upholstery, such as armrest pads, calf pads, headrest or neckrest, should be cleaned according to the instruction above.

8.3.5 Disinfection

Ĭ

The wheelchair may be disinfected by spraying or wiping with tested, approved disinfectants.

- Spray a soft cleaning and disinfecting product (bactericidal and fungicide meeting the EN1040/EN1276 / EN1650 standards) and follow the instructions given by the manufacturer.
- 1. Wipe down all generally accessible surfaces with a soft cloth and ordinary household disinfectant.
- 2. Allow the product to air-dry.

9 Troubleshooting

9.1 Safety Information

Faults may arise as a result of daily use, adjustments or changing demands on the wheelchair. The table below shows how to identify and repair faults. Some of the actions listed must be carried out by a qualified technician. These are indicated. We recommend that all adjustments are carried out by a qualified technician.



CAUTION!

 If you notice a fault with your wheelchair, e.g. a significant change in handling, stop using your wheelchair immediately and contact your provider.

30 1608467-1

9.2 Identifying and repairing faults

Fault Possible cause Action			
rauit		Action	
	Incorrect tyre pressure on one rear wheel	Correct tyre pressure, → 11.3 Tyres, page 33	
The wheelchair does not travel in	One or more spokes broken	Replace faulty spoke(s), → qualified technician	
a straight line	Spokes tightened unevenly	Tighten loose spokes, → qualified technician	
	Castor bearings are dirty or damaged	Clean the bearings or replace the castor, → qualified technician	
	Rear wheels are mounted too far forwards	Mount the rear wheels further back, → qualified technician	
The wheelchair tips backwards	Back angle too large	Reduce the backrest angle, → qualified technician	
too easily	Seat angle too large	Mount the castor upper on the front fork, → qualified technician Select a smaller front fork, → qualified technician	
The brakes are gripping poorly or	Incorrect tyre pressure in one or both rear tyres	Correct tyre pressure, → 11.3 Tyres, page 33	
asymmetrically	Brake setting incorrect	Correct the brake setting, → qualified technician	
Rolling resistance	Tyre pressure in rear tyres is too low	Correct tyre pressure, → 11.3 Tyres, page 33	
is very high	Rear wheels not parallel	Ensure the rear wheels are parallel, → qualified technician	
The castors wobble when	Too little tension in castor bearing block	Tighten the nut on the bearing block axle slightly, → qualified technician	
moving fast	Castor has worn smooth	Change castor, → qualified technician	
The castor is stiff	Bearings are	Clean the bearings	

Fault	Possible cause	Action
or stuck	dirty or faulty	or replace the castor, → qualified technician
The wheelchair is very difficult to unfold	Backrest cover is too tight	Loosen the hook and loop strips of the backrest cover a little, → 3.5.2 Adjustable Backrest Cover, page 7

10 After use

10.1 Storage

NOTICE!

Risk of damage to the product

Do not store the product near heat sources. Never store other items on top of the wheelchair. Store the wheelchair indoors in a dry environment. Refer to temperature limitation in section *11* Technical data, page 32.

After long-term storage (more than four months) the wheelchair must be inspected in accordance to section 8.2 Maintenance schedule, page 28.

10.2 Disposal

Be environmentally responsible and recycle this product through your recycling facility at its end of life.

Disassemble the product and its components, so the different materials can be separated and recycled individually.

The disposal and recycling of used products and packaging must comply with the laws and regulations for waste handling in each country. Contact your local waste management company for information.

10.3 Reconditioning

This product is suitable for reuse. To recondition the product for a new user, carry out the following actions:

- Inspection
- · Cleaning and disinfection
- Adaptation to the new user

For detailed information, please see section 8.2 Maintenance schedule, page 28 and the service manual for this product. Make sure that the user manual is handed over with the product.

If any damage or malfunction is detected, do not reuse the product.


11 Technical data


11.1 Dimensions and weight

All dimension and weight specifications refer to a wide range of the wheelchair in a standard configuration. Dimension and weight (based on ISO 7176–1/5/7) may alter according to different configurations. In some configurations, the wheelchair exceeds the recommended dimensions for a wheelchair.

NOTICE!

- In some configurations, the overall dimensions of the wheelchair when it is ready for use exceeds the authorised limits and the access to emergency escape routes is not possible.
- In some configurations, the wheelchair exceeds the size recommended for travelling by train in the EU.





		Action 3 NG
	Maximum load of the wheelchair (*)	125 kg
A	Overall length with leg rest	962 – 1185 mm
B	Overall width	570 – 710 mm
	Folded length	765 – 1200 mm
	Folded width	285 – 355 mm
	Folded height	795 - 1510 mm
	Total mass	from 12,7 – 28,2 kg
	Mass of the heaviest part	7,5 – 12,4 kg
	Static stability	Downhill: 0° – 17° Uphill: 0° – 7° Sideways: 0° – 17°
	Seat plane angle	0° - 8° (0°/3°/6°)
©	Effective seat depth	395 – 500
D	Effective seat width	370 – 505 mm, in increments of 25 mm
Ē	Seat surface height at front edge	345 - 550 mm, in increments of 25 mm
Ē	Seat surface height at rear edge	345 - 550 mm, in increments of 25 mm

		Action 3 NG
©	Backrest angle	Range: -15° / +30° Fixed: 90° Adjustable: 90° (- 15°/+15°), step less Recliner: 90° (0°/+30°), in increments of 10°
Θ	Backrest height	325 - 610 mm
①	Footrest to seat distance	160 - 565 mm, in increments of 10 mm
	Leg to seat surface angle	0° – 90°
1	Armrest to seat height	Range: 170 – 300mm N°1: 200 – 280 mm N°2: 240 mm N°3: 170 – 295 mm N°4: 210 – 300 mm
	Front location of armrest structure	190 – 350 mm
	Hand rim diameter	520 – 540 mm
	Horizontal location of axle	–55 / + 42,5 mm
	Minimum turning radius	790 mm
	Stowage width	250 – 320 mm
	Stowage height	566 – 1000 mm
	Stowage length	from 695 – 1130 mm
K	Overall length without leg rests	720 – 860 mm
(L)	Overall height	795 – 1510 mm
	Pivot width	1120 – 1505 mm
	Maximum slope angle brake	7°

11.2 Maximum Weight of Removable Parts

Part	Maximum weight
Leg rest angle adjustable with calf pad and footrest	3,2 kg
Armrest	2,0 kg
Rear wheel 600 mm (24") solid with handrim and spoke guard	2,5 kg
Headrest / Neckrest / Cheek support	1,4 kg
Trunk support	0,8 kg
Backrest cushion	2,0 kg
Seat cushion	1,9 kg
Table tray	3,9 kg

32 1608467-I

11.3 Tyres

The ideal pressure depends on the tyre type:

Tyre	Max. pressure		
Pneumatic profiled tyre: 515 mm (20"); 560 mm (22"); 610 mm (24")	4.5 bar	450 kPa	65 psi
Pneumatic low profile tyre: 560 mm (22"); 610 mm (24")	7.5 bar	750 kPa	110 psi
Schwalbe® Marathon Plus Evolution tyre: 610 mm (24")	10 bar	1000 kPa	145 psi
Solid tyre: 305 mm (12"); 515 mm (20"); 560 mm (22"); 610 mm (24")	-	-	-
Solid tyre: 150 mm (6"); 140 mm (5" 1/2); 125 mm (5"); 200 mm (8")	-	-	-
Pneumatic tyre: 200 mm (8")	2.5 bar	250 kPa	36 psi

- The compatibility of the tyres listed above depends on the configuration and/or model of your wheelchair.
- In case of a tyre puncture consult a suitable workshop (e.g. bike repair shop, bicycle dealer ...) to have the tube replaced by a skilled person.
- The size of the tyre is mentioned on the sidewall of the tyre. The change of appropriate tyres must be carried out by a qualified technician.



CAUTION!

 The tyre pressure has to be equal in both wheels to avoid a less driving comfort, to keep the brakes efficiency and an easy propelling of the wheelchair.

11.4 Materials

Part	Material
Frame / backrest tubes	Aluminum, Steel
Upholstery (seat and backrest)	Foam PUR, Fabric Nylon and Polyester
Push handles	PVC
Plastic parts like brake handles, clothes guards, foot plates, armpads and parts of most options	Thermoplastic (e.g. PA, PP, ABS and PUR) according to marking on the parts
Folding mechanism / vertical brace / clamping parts / castor forks	Aluminum, Steel
Screws, washers and nuts	Steel

- All materials used are protected against corrosion. We use only REACH compliant materials and components.
- Theft and metal detection systems: in seldom cases the materials used in the wheelchair may activate theft and metal detection systems.

11.5 Environmental conditions

	Storage and transportation	Operation
Temperature	-20 °C to 40 °C	-5 °C to 40 °C
Relative humidity	20 % to 90 % at 30 °C,	not condensing
Atmospheric pressure	800 hPa to 1060 hPa	

Be aware that when a wheelchair has been stored under low temperatures, it must be adjusted before use. Please see 8.1 Safety Information, page 28.



United Kingdom:

Invacare Limited Pencoed Technology Park, Pencoed Bridgend CF35 5AQ Tel: (44) (0) 1656 776 200 Tel: (44) (0) 1656 776 200 uk@invacare.com

New Zealand:

www.invacare.co.uk

Invacare New Zealand Ltd 4 Westfield Place, Mt Wellington 1060 New Zealand

Phone: 0800 468 222 Fax: 0800 807 788 sales@invacare.co.nz

Ireland:

Invacare Ireland Ltd, Unit 5 Seatown Business Campus Seatown Road, Swords, County Dublin EirCode: K67 K271 — Ireland

Tel: (353) 1 810 7084 Fax: (353) 1 810 7085 ordersireland@invacare.com

Asia:

Invacare Asia Ltd. Unit 18/12 Stanton Road, Seven Hills, NSW 2147 Australia Phone: (61) (02) 8839 5333

Fax: (61) (02) 8839 5343 asiasales@invacare.com

Australia:

Invacare Australia Pty. Ltd. Building 5, Unit F, 161 Manchester Road, Auburn, Sydney, NSW 2144 Australia

Phone: 1800 460 460 Fax: 1800 814 367 orders@invacare.com.au www.invacare.com.au

CE K



Invacare UK Operations Limited UKRP Unit 4, Pencoed Technology Park, Pencoed Bridgend CF35 5AQ UK



Invacare France Operations SAS Route de St Roch F-37230 Fondettes France

1608467-I 2025-11-24



